

# Prüfprotokoll OGM

Oval gear meter type 0.1

PN: 280001

OGM 001VCFPKKEI-G1/8i-99-2600-0003

Materials of construction:

PVC, FPM, PEEK,  
ceramic, titan

Signal output:

2600 pulse/l

Max. flow rate:

3 l/h

Prod. code:

494239 / 41422 / 1001

ECOLAB Engineering GmbH

D-83313 Siegsdorf



Anlaufgrenze 0,80 [l/h]\*  
lower measuring range limit

Impulszahl 55 [n/min]  
pulse rate

bei 1,25 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 19.4.22

Prüfer P12

# Prüfprotokoll OGM

Oval gear meter type 0.1

PN: 280001

OGM 001VCFPPKKEI-G1/8i-99-2600-0003

Materials of construction:

PVC, FPM, PEEK,  
ceramic, titan

Signal output:

2600 pulse/l

Max. flow rate:

3 l/h

Prod. code:

494239 / 4.14.22 / 1002

ECOLAB Engineering GmbH

D-83313 Siegedorf



Anlaufgrenze 0,80 [l/h]\*  
lower measuring range limit

Impulszahl 55 [n/min]  
pulse rate

bei 1,25 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 19.4.22

Prüfer P12

# Prüfprotokoll OGM

Oval gear meter type 0.1

PN: 280001

OGM 001VCFPPKKEI-G1/8i-99-2600-0003

Materials of construction:

PVC, FPM, PEEK,  
ceramic, titan  
2800 pulse/l

Signal output:

Max. flow rate:

Prod. code:

3 l/h  
494239 / 41422 / 1003

ECOLAB Engineering GmbH  
D-83313 Siegedorf



Anlaufgrenze 0,80 [l/h]\*  
lower measuring range limit

Impulszahl 56 [n/min]  
pulse rate

bei 1,25 [l/h]\*  
at


\* = kontinuierlicher Durchfluss / continuous flow

Datum 19.4.22

Prüfer P12

# Prüfprotokoll OGM

Oval gear meter type 0.1  
PN: 280001  
OGM 001VCFPPKKEI-G1/8I-99-2600-0003  
Materials of construction: PVC, FPM, PEEK, ceramic, titan  
Signal output: 2600 pulse/l  
Max. flow rate: 3 l/h  
Prod. code: 494239 / 414 22 / 1004  
ECOLAB Engineering GmbH  
D-83313 Siegsdorf



Anlaufgrenze 0,80 [l/h]\*  
lower measuring range limit

Impulszahl 53 [n/min]  
pulse rate

bei 1,25 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 19.4.22

Prüfer P12

# Prüfprotokoll OGM

Oval gear meter type 0.1

PN: 250001

OGM 001VCFPPKKEI-G1/8i-99-2600-0003

Materials of construction:

PVC, FPM, PEEK,  
ceramic, titan

Signal output:

2600 pulse/l

Max. flow rate:

3 l/h

Prod. code:

494239 / 41422 / 1005

ECOLAB Engineering GmbH  
D-83313 Siegedorf



Anlaufgrenze 0,80 [l/h]\*  
lower measuring range limit

Impulszahl 55 [n/min]  
pulse rate

bei 1,25 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 19.4.22

Prüfer P12

# Prüfprotokoll OGM

Oval gear meter type 0.1

PN: 280001

OGM 001VCFPKKEI-G1/8i-99-2600-0003

Materials of construction:

PVC, FPM, PEEK,  
ceramic, titan  
2800 pulse/l

Signal output:

Max. flow rate:

Prod. code:

494239 / 41422 / 1006

ECOLAB Engineering GmbH  
D-83313 Siegsdorf



Anlaufgrenze 0,80 [l/h]\*  
lower measuring range limit

Impulszahl 52 [n/min]  
pulse rate

bei 1,25 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 19.4.22

Prüfer P12