

Prüfprotokoll OGM

| | | |
|--|-----------------------------------|---|
| Oval gear meter type 1 | |  |
| PN: 280002 | | |
| OGM 010VCFPKKEI-G1/4i-99-0820-0030 | | |
| Materials of construction: | PVC, FPM, PEEK, ceramic, titan | |
| Signal output: | 820 pulse/l | |
| Max. flow rate: | 90 l/h | |
| Prod. code: | 495535 / 21522 / 1001 | |
| ECOLAB Engineering GmbH D-83313 Siegsdorf | | |

Anlaufgrenze 5,5 [l/h]*
lower measuring range limit

Impulszahl 125 [n/min]
pulse rate

bei 9,0 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 19.4.22

Prüfer P12

Prüfprotokoll OGM

Oval gear meter type 1

PN: 280002

OGM 010VCFPPKETI-G1/4i-99-0820-0030

Materials of construction:

PVC, FPM, PEEK,
ceramic, titan

Signal output:

820 pulse/l

Max. flow rate:

30 l/h

Prod. code:

495535 / 21522/1002

ECOLAB Engineering GmbH

D-83313 Siegedorf



Anlaufgrenze 5,5 [l/h]*
lower measuring range limit

Impulszahl 126 [n/min]
pulse rate

bei 9,0 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 19.4.22

Prüfer P12