



Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 02100		
PN: 280065		
OGM PLUS 02100VCFPPKKE-G5/4I-99-0125p-P		
materials of construction:	PVC, FPM, PEEK, ceramic	
signal output (standard):	8,0 ml/pulse, 125 pulses/l	
max. pump size:	210 / 250 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	704946/13822/1001	
ECOLAB Engineering GmbH D-83313 Siegsdorf		

Anlaufgrenze 50 [l/h]*
lower measuring range limit

Impulszahl 183 [n/min]
pulse rate

bei 80 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 22.09.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 02100

PN: **280065**



OGM PLUS 02100VCFPPKKE-G5/4I-99-0125p-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 8,0 ml/pulse, 125 pulses/l

max. pump size: 210 / 250 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 704948/13822/1002

ECOLAB Engineering GmbH
D-83313 Siegsdorf



Anlaufgrenze 50 [l/h]*
lower measuring range limit

Impulszahl 191 [n/min]
pulse rate



bei 80 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 22.09.22

Prüfer **P 410**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 02100		
PN: 280065		
OGM PLUS 02100VCFPPKKE-G5/4I-99-0125p-P		
materials of construction:	PVC, FPM, PEEK, ceramic	
signal output (standard):	8,0 ml/pulse, 125 pulses/l	
max. pump size:	210 / 250 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	704948 / 13822 / 1003	
ECOLAB Engineering GmbH D-83313 Siegsdorf		

Anlaufgrenze 50 [l/h]*
lower measuring range limit

Impulszahl 184 [n/min]
pulse rate

bei 80 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 22.09.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 02100

PN: **280065**



OGM PLUS 02100VCPPKKE-G5/4I-99-0125p-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 8,0 ml/pulse, 125 pulses/l

max. pump size: 210 / 250 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 704948 / 13822 / 1004

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 50 [l/h]*
lower measuring range limit

Impulszahl 186 [n/min]
pulse rate

bei 80 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 22.09.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 02100

PN: **280065**



OGM PLUS 02100VCFPPKKE-G5/4I-99-0125p-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 8,0 ml/pulse, 125 pulses/l

max. pump size: 210 / 250 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 704948 /13822 /1.005

ECOLAB Engineering GmbH
D-83313 Siegsdorf



Anlaufgrenze 50 [l/h]*
lower measuring range limit

Impulszahl 191 [n/min]
pulse rate

bei 80 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 22.09.22

Prüfer P 410