

Prüfprotokoll OGM

Oval gear meter type 1

PN: 280002

OGM 010VCFPPKKEI-G1/4i-99-0820-0030

Materials of construction:

PVC, FPM, PEEK,
ceramic, titan

Signal output:

820 pulse/l

Max. flow rate:

30 l/h

Prod. code:

704344 / 343 22 / 1002

ECOLAB Engineering GmbH

D-83313 Siegedorf



Anlaufgrenze 5,5 [l/h]*
lower measuring range limit

Impulszahl 118 [n/min]
pulse rate

bei 9,0 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28.10.22

Prüfer P391

Prüfprotokoll OGM

Oval gear meter type 1

PN: 280002

OGM 010VCFPPKETI-G1/4i-99-0820-0030

Materials of construction:

PVC, FPM, PEEK,
ceramic, titan

Signal output:

820 pulse/l

Max. flow rate:

30 l/h

Prod. code:

704344 / 34322 / 1001

ECOLAB Engineering GmbH

D-83313 Siegsdorf



Anlaufgrenze 5,5 [l/h]*
lower measuring range limit

Impulszahl 120 [n/min]
pulse rate

bei 9,0 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28.10.22

Prüfer P391