

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200

PN: 10240555



OGM PLUS 01200VCFPPKKE-G3/4I-99-0005m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 5.0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

Year/Prod. code: 2022 / 705109/14822/1001

ECOLAB Engineering GmbH

D-83313 Slegsdorf

Made in Germany



Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 108 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 29.11.22

Prüfer Sollereder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200

PN: **10240555**



OGM PLUS 01200VCFPPKKE-G3/4I-99-0005m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 5,0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

Year/Prod. code: 2022 / 705109 / 14822 / 1002

ECOLAB Engineering GmbH
D-83313 Siegsdorf
Made in Germany



Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 110 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 29.11.22

Prüfer Sollereder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200		
PN: 10240555		
OGM PLUS 01200VCFPPKKE-G3/4I-99-0005m-P		
materials of construction:	PVC, FPM, PEEK, ceramic	
signal output (standard):	5,0 ml/pulse, 200 pulses/l	
max. pump size:	120 / 144 l/h (50/60Hz)	
semiconductor junction:	PNP	
Year/Prod. code:	2022 / 705109 / 14822 / 1003	
ECOLAB Engineering GmbH D-83313 Slegsdorf Made in Germany		

Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 114 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 29.11.22

Prüfer Sillereeder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200

PN: **10240555**



OGM PLUS 01200VCFPPKKE-G3/4I-99-0005m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 5,0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

Year/Prod. code: 2022 / 705109 / 1 48 22 / 1004

ECOLAB Engineering GmbH

D-83313 Siegedorf

Made in Germany



Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 106 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 29.11.22

Prüfer Sollereder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200		
PN: 10240555		
OGM PLUS 01200VCFPPKKE-G3/4I-99-0005m-P		
materials of construction:	PVC, FPM, PEEK ceramic	
signal output (standard):	5.0 ml/pulse, 200 pulses/l	
max. pump size:	120 / 144 l/h (50/60Hz)	
semiconductor junction:	PNP	
Year/Prod. code:	2022 / 705109 / 14822 / 1005	
ECOLAB Engineering GmbH D-83313 Slegendorf Made in Germany		

Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 107 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 29.11.22

Prüfer Sollereder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200

PN: 10240555



OGM PLUS 01200VCFPPKKE-G3/4I-99-0005m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 5.0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

Year/Prod. code: 2022 / 705109 / 1 48 22 / 1 006

ECOLAB Engineering GmbH

D-83313 Siegsdorf

Made in Germany



Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 114 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 29.11.22

Prüfer Sollereder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200

PN: 10240555

OGM PLUS 01200VCFPPKKE-G3/4I-99-0005m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 5.0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

Year/Prod. code: 2022 / 705109 / 14822 / 1007

ECOLAB Engineering GmbH

D-83313 Siedsdorf

Made in Germany



Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 108 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 29.11.22

Prüfer Sollwender

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200

PN: **10240555**



OGM PLUS 01200VCFPPKE-G3/4I-99-0005m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 5.0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

Year/Prod. code: 2022 / 705109 / 148 22 / 1008

ECOLAB Engineering GmbH
D-83313 Siegsdorf
Made in Germany



Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 114 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 29.11.22

Prüfer Sollereder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200

PN: **10240555**

OGM PLUS 01200VCFPPKKE-G3/4I-99-0005m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 5,0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

Year/Prod. code: 2022 / 705109 / 14822 / 1009

ECOLAB Engineering GmbH
D-83313 Siegsdorf
Made in Germany



Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 115 [n/min]
pulse rate



bei 55 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 29.11.22

Prüfer Sollereder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200		
PN: 10240555		
OGM PLUS 01200VCFPPKKE-G3/4I-99-0005m-P		
materials of construction:	PVC, FPM, PEEK, ceramic	
signal output (standard):	5.0 ml/pulse, 200 pulses/l	
max. pump size:	120 / 144 l/h (50/60Hz)	
semiconductor junction:	PNP	
Year/Prod. code:	2022 / 705109 / 1 48 22 / 1010	
ECOLAB Engineering GmbH D-83313 Slegsdorf Made in Germany		

Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 111 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 29.11.22

Prüfer Sollwieder