

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200 
PN: **W3T267463**
OGM PLUS 01200VCFPPKKE-G3/4I-99-0300p-P
materials of construction: PVC, FPM, PEEK, ceramic
signal output (standard): 3,33 ml/pulse, 300 pulses/l
max. pump size: 120 / 144 l/h (50/60Hz)
semiconductor junction: PNP
prod. code: 01 / 708363 / 40823 / 1001
ECOLAB Engineering GmbH
D-83313 Siegedorf 

Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 168 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28.02.2023

Prüfer P 421

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200		
PN: W3T267463		
OGM PLUS 01200VCFPPKKE-G3/4I-99-0300p-P		
materials of construction:	PVC, FPM, PEEK, ceramic	
signal output (standard):	3,33 ml/pulse, 300 pulses/l	
max. pump size:	120 / 144 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	01/708363/40823/1002	
ECOLAB Engineering GmbH D-63313 Siegedorf		

Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 165 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28.02.2023

Prüfer P 421

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200 
PN: **W3T267463**
OGM PLUS 01200VCFPPKKE-G3/4I-99-0300p-P
materials of construction: PVC, FPM, PEEK, ceramic
signal output (standard): 3,33 ml/pulse, 300 pulses/l
max. pump size: 120 / 144 l/h (60/60Hz)
semiconductor junction: PNP
prod. code: 01/708363 / 40823 / 1003
ECOLAB Engineering GmbH
D-83313 Siegedorf 

Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 170 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28.02.2023

Prüfer P. 421