



Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00540		
PN: 280073		
OGM PLUS 00540VCEPPKKE-G1/4I-99-0005m-P		
materials of construction:	PVC, EPDM, PEEK, ceramic	
signal output (standard):	5,0 ml/pulse, 200 pulses/l	
max. pump size:	54 / 85 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	01 / 497831 / 32822 / 1001	
ECOLAB Engineering GmbH D-83313 Siegsdorf		

Anlaufgrenze 6 [l/h]*
lower measuring range limit

Impulszahl 34 [n/min]
pulse rate



bei 10 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 29.07.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00540		
PN: 280073		
OGM PLUS 00540VCEPPKKE-G1/4I-99-0005m-P		
materials of construction:	PVC, EPDM, PEEK, ceramic	
signal output (standard):	5,0 ml/pulse, 200 pulses/l	
max. pump size:	54 / 65 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	01 / 497831 / 32822 / 1002	
ECOLAB Engineering GmbH D-83313 Siegedorf		

Anlaufgrenze 6 [l/h]*
lower measuring range limit

Impulszahl 34 [n/min]
pulse rate

bei 10 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 29.07.22

Prüfer **P 410**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00540

PN: **280073**

OGM PLUS 00540VCEPPKKE-G1/4I-99-0005m-P

materials of construction: PVC, EPDM, PEEK, ceramic

signal output (standard): 5,0 ml/pulse, 200 pulses/l

max. pump size: 54 / 65 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/497831/32822/1003

ECOLAB Engineering GmbH

D-83313 Siegsdorf



Anlaufgrenze 6 [l/h]*
lower measuring range limit

Impulszahl 36 [n/min]
pulse rate



bei 10 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 29.07.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00540		
PN: 280073		
OGM PLUS 00540VCEPPKKE-G1/4I-99-0005m-P		
materials of construction:	PVC, EPDM, PEEK, ceramic	
signal output (standard):	5.0 ml/pulse, 200 pulses/l	
max. pump size:	54 / 65 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	01 / 497831 / 32822 / 1004	
ECOLAB Engineering GmbH D-83313 Siegsdorf		

Anlaufgrenze 6 [l/h]*
lower measuring range limit

Impulszahl 35 [n/min]
pulse rate



bei 10 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 29.07.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00540	
PN: 290073	
OGM PLUS 00540VCEPPKKE-G1/4I-99-0005m-P	
materials of construction:	PVC, EPDM, PEEK, ceramic
signal output (standard):	5.0 ml/pulse, 200 pulses/l
max. pump size:	54 / 65 l/h (50/60Hz)
semiconductor junction:	PNP
prod. code:	01/497831/32822/1005
ECOLAB Engineering GmbH D-83313 Siegedorf	

Anlaufgrenze 6 [l/h]*
lower measuring range limit

Impulszahl 32 [n/min]
pulse rate

bei 10 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 29.07.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00540

PN: **280073**

OGM PLUS 00540VCEPPKKE-G1/4I-99-0005m-P

materials of construction: PVC, EPDM, PEEK, ceramic

signal output (standard): 5.0 ml/pulse, 200 pulses/l

max. pump size: 54 / 65 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/497831/32822/1006

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 6 [l/h]*
lower measuring range limit

Impulszahl 34 [n/min]
pulse rate


bei 10 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 29.07.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00540		
PN: 280073		
OGM PLUS 00540VCEPPKKE-G1/4I-00-0005m-P		
materials of construction:	PVC, EPDM, PEEK, ceramic	
signal output (standard):	5,0 ml/pulse, 200 pulses/l	
max. pump size:	54 / 65 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	01/497831/32822/1007	
ECOLAB Engineering GmbH D-83313 Siegedorf		

Anlaufgrenze 6 [l/h]*
lower measuring range limit

Impulszahl 34 [n/min]
pulse rate



bei 10 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 29.07.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00540		
PN: 260073		
OGM PLUS 00540VCEPPKKE-G1/4I-99-0005m-P		
materials of construction:	PVC, EPDM, PEEK, ceramic	
signal output (standard):	5,0 ml/pulse, 200 pulses/l	
max. pump size:	54 / 65 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	01/497831/32822/1008	
ECOLAB Engineering GmbH D-83313 Siegsdorf		

Anlaufgrenze 6 [l/h]*
lower measuring range limit

Impulszahl 33 [n/min]
pulse rate

bei 10 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 29.07.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00540

PN: **280073**

OGM PLUS 00540VCEPPKKE-G1/4I-99-0005m-P

materials of construction: PVC, EPDM, PEEK, ceramic

signal output (standard): 5,0 ml/pulse, 200 pulses/l

max. pump size: 54 / 85 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01 / 497831 / 32822 / 1009

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 6 [l/h]*
lower measuring range limit

Impulszahl 35 [n/min]
pulse rate

bei 10 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 29.07.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00540

PN: **280073**

OGM PLUS 00540VCEPPKKE-G1/4I-99-0005m-P

materials of construction: PVC, EPDM, PEEK, ceramic

signal output (standard): 5,0 ml/pulse, 200 pulses/l

max. pump size: 54 / 65 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/497831/32822/1010

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 6 [l/h]*
lower measuring range limit

Impulszahl 34 [n/min]
pulse rate

bei 10 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 29.07.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00540	
PN: 280073	
OGM PLUS 00540VCEPPKKE-G1/4I-99-0005m-P	
materials of construction:	PVC, EPDM, PEEK, ceramic
signal output (standard):	5,0 ml/pulse, 200 pulses/l
max. pump size:	54 / 65 l/h (50/60Hz)
semiconductor junction:	PNP
prod. code:	01/497831/32822/1011
ECOLAB Engineering GmbH D-83313 Siegedorf	

Anlaufgrenze 6 [l/h]*
lower measuring range limit

Impulszahl 31 [n/min]
pulse rate



bei 10 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 29.07.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00540		
PN: 280073		
OGM PLUS 00540VCEPPKKE-G1/4I-99-0005m-P		
materials of construction:	PVC, EPDM, PEEK, ceramic	
signal output (standard):	5,0 ml/pulse, 200 pulses/l	
max. pump size:	54 / 65 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	01 / 497831 / 32822 / 1012	
ECOLAB Engineering GmbH D-63313 Siegedorf		

Anlaufgrenze 6 [l/h]*
lower measuring range limit

Impulszahl 33 [n/min]
pulse rate



bei 10 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 29.07.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00540	
PN: 280073	
OGM PLUS 00540VCEPPKKE-G1/4I-99-0005m-P	
materials of construction:	PVC, EPDM, PEEK, ceramic
signal output (standard):	5,0 ml/pulse, 200 pulses/l
max. pump size:	54 / 65 l/h (50/60Hz)
semiconductor junction:	PNP
prod. code:	01/497831 /32822 /1013
ECOLAB Engineering GmbH D-83313 Siegedorf	

Anlaufgrenze 6 [l/h]*
lower measuring range limit

Impulszahl 35 [n/min]
pulse rate

bei 10 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 29.07.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00540
PN: **280073**
OGM PLUS 00540VCEPPKKE-G1/4I-99-0005m-P
materials of construction: PVC, EPDM, PEEK, ceramic
signal output (standard): 5,0 ml/pulse, 200 pulses/l
max. pump size: 54 / 65 l/h (50/60Hz)
semiconductor junction: PNP
prod. code: 01/497831/32822/1014

ECOLAB Engineering GmbH
D-83313 Siegedorf

Anlaufgrenze 6 [l/h]*
lower measuring range limit

Impulszahl 36 [n/min]
pulse rate



bei 10 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 29.07.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00540		
PN: 280073		
OGM PLUS 00540VCEPPKKE-G1/4I-99-0005m-P		
materials of construction:	PVC, EPDM, PEEK, ceramic	
signal output (standard):	5,0 ml/pulse, 200 pulses/l	
max. pump size:	54 / 85 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	01/497831 /32822/1015	
ECOLAB Engineering GmbH D-83313 Slegedorf		

Anlaufgrenze 6 [l/h]*
lower measuring range limit

Impulszahl 33 [n/min]
pulse rate

bei 16 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 29.07.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00540		
PN: 280073		
OGM PLUS 00540VCEPPKKE-G1/4I-99-0005m-P		
materials of construction:	PVC, EPDM, PEEK, ceramic	
signal output (standard):	5,0 ml/pulse, 200 pulses/l	
max. pump size:	54 / 85 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	01/497831/32822/1018	
ECOLAB Engineering GmbH D-83313 Siegsdorf		

Anlaufgrenze 6 [l/h]*
lower measuring range limit

Impulszahl 31 [n/min]
pulse rate



bei 10 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 29.07.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00540		
PN: 280073		
OGM PLUS 00540VCEPPKKE-G1/4I-99-0005m-P		
materials of construction:	PVC, EPDM, PEEK, ceramic	
signal output (standard):	5,0 ml/pulse, 200 pulses/l	
max. pump size:	54 / 85 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	01/497831 /32822 /1017	
ECOLAB Engineering GmbH D-83313 Blegdorf		

Anlaufgrenze 6 [l/h]*
lower measuring range limit

Impulszahl 35 [n/min]
pulse rate

bei 10 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 29.07.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00540		
PN: 280073		
OGM PLUS 00540VCEPPKKE-G1/4I-99-0005m-P		
materials of construction:	PVC, EPDM, PEEK, ceramic	
signal output (standard):	5,0 ml/pulse, 200 pulses/l	
max. pump size:	54 / 85 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	01/497831/32822/1018	
ECOLAB Engineering GmbH D-83313 Siegedorf		

Anlaufgrenze 6 [l/h]*
lower measuring range limit

Impulszahl 34 [n/min]
pulse rate



bei 10 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 29.07.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00540	
PN: 280073	
OGM PLUS 00540VCEPPKKE-G1/4I-99-0005m-P	
materials of construction:	PVC, EPDM, PEEK, ceramic
signal output (standard):	5,0 ml/pulse, 200 pulses/l
max. pump size:	54 / 85 l/h (50/60Hz)
semiconductor junction:	PNP
prod. code:	01/497831/32622/1019
ECOLAB Engineering GmbH	
D-83313 Siegedorf	

Anlaufgrenze 6 [l/h]*
lower measuring range limit

Impulszahl 35 [n/min]
pulse rate

bei 10 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 29.07.22

Prüfer **P 410**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00540

PN: **280073**

OGM PLUS 00540VCEPPKKE-G1/4I-99-0005m-P

materials of construction: PVC, EPDM, PEEK, ceramic

signal output (standard): 5,0 ml/pulse, 200 pulses/l

max. pump size: 54 / 85 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/497831/32822/1020

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 6 [l/h]*
lower measuring range limit

Impulszahl 34 [n/min]
pulse rate



bei 10 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 29.07.22

Prüfer **P 410**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00540		
PN: 280073		
OGM PLUS 00540VCEPPKKE-G1/4I-99-0005m-P		
materials of construction:	PVC, EPDM, PEEK, ceramic	
signal output (standard):	5.0 ml/pulse, 200 pulses/l	
max. pump size:	54 / 65 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	01 / 497831 / 32822 / 1021	
ECOLAB Engineering GmbH D-83113 Siegedorf		

Anlaufgrenze 6 [l/h]*
lower measuring range limit

Impulszahl 32 [n/min]
pulse rate



bei 10 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 29.07.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00540		
PN: 280073		
OGM PLUS 00540VCEPPKKE-G1/4I-99-0005m-P		
materials of construction:	PVC, EPDM, PEEK, ceramic	
signal output (standard):	5.0 ml/pulse, 200 pulses/l	
max. pump size:	54 / 65 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	01/497831 /32822 /1022	
ECOLAB Engineering GmbH D-83313 Siegsdorf		

Anlaufgrenze 6 [l/h]*
lower measuring range limit

Impulszahl 34 [n/min]
pulse rate

bei 10 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 29.07.22

Prüfer **P 410**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00540		
PN: 280073		
OGM PLUS 00540VCEPPKKE-G1/4I-99-0005m-P		
materials of construction:	PVC, EPDM, PEEK, ceramic	
signal output (standard):	5.0 ml/pulse, 200 pulses/l	
max. pump size:	54 / 65 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	01/497831/32822/1023	
ECOLAB Engineering GmbH D-83313 Siegsdorf		

Anlaufgrenze 6 [l/h]*
lower measuring range limit

Impulszahl 35 [n/min]
pulse rate

bei 10 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 29.07.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00540

PN: **280073**

OGM PLUS 00540VCEPPKKE-G1/4I-99-0005m-P

materials of construction: PVC, EPDM, PEEK, ceramic

signal output (standard): 5.0 ml/pulse, 200 pulses/l

max. pump size: 54 / 65 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/497831/32822/1024

ECOLAB Engineering GmbH
D-83313 Stegedorf



Anlaufgrenze 6 [l/h]*
lower measuring range limit

Impulszahl 35 [n/min]
pulse rate



bei 10 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 29.07.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00540		
PN: 280073		
OGM PLUS 00540VCEPPKKE-G1/4I-99-0005m-P		
materials of construction:	PVC, EPDM, PEEK, ceramic	
signal output (standard):	5,0 ml/pulse, 200 pulses/l	
max. pump size:	54 / 65 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	01/497631 /32822 /1025	
ECOLAB Engineering GmbH D-83313 Stegedorf		

Anlaufgrenze 6 [l/h]*
lower measuring range limit

Impulszahl 35 [n/min]
pulse rate

bei 10 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 29.07.22

Prüfer **P 410**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00540

PN: **280073**

OGM PLUS 00540VCEPPKKE-G1/4I-99-0005m-P

materials of construction: PVC, EPDM, PEEK, ceramic

signal output (standard): 5.0 ml/pulse, 200 pulses/l

max. pump size: 54 / 65 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/497831/32822/1026

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 6 [l/h]*
lower measuring range limit

Impulszahl 33 [n/min]
pulse rate



bei 10 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 29.07.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00540	
PN: 280073	
OGM PLUS 00540VCEPPKKE-G1/4I-99-0005m-P	
materials of construction:	PVC, EPDM, PEEK, ceramic
signal output (standard):	5.0 ml/pulse, 200 pulses/l
max. pump size:	54 / 65 l/h (50/60Hz)
semiconductor junction:	PNP
prod. code:	01/497831/32822/1027
ECOLAB Engineering GmbH D-83313 Siegedorf	

Anlaufgrenze 6 [l/h]*
lower measuring range limit

Impulszahl 36 [n/min]
pulse rate


bei 10 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 29.07.22

Prüfer **P 410**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00540		
PN: 280073		
OGM PLUS 00540VCEPPKKE-G1/4I-99-0005m-P		
materials of construction:	PVC, EPDM, PEEK, ceramic	
signal output (standard):	5.0 ml/pulse, 200 pulses/l	
max. pump size:	54 / 65 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	01/497831 /32822 /1028	
ECCOLAB Engineering GmbH D-83313 Slegsdorf		

Anlaufgrenze 6 [l/h]*
lower measuring range limit

Impulszahl 35 [n/min]
pulse rate



bei 10 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 29.07.22

Prüfer **P 410**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00540	
PN: 280073	
OGM PLUS 00540VCEPPKKE-G1/4I-99-0005m-P	
materials of construction:	PVC, EPDM, PEEK, ceramic
signal output (standard):	5,0 ml/pulse, 200 pulses/l
max. pump size:	54 / 65 l/h (50/60Hz)
semiconductor junction:	PNP
prod. code:	01/497831/32822/1029
ECOLAB Engineering GmbH D-83313 Siegedorf	

Anlaufgrenze 6 [l/h]*
lower measuring range limit

Impulszahl 34 [n/min]
pulse rate

bei 10 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 29.07.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00540

PN: **280073**

OGM PLUS 00540VCEPPKKE-G1/4I-99-0005m-P

materials of construction: PVC, EPDM, PEEK, ceramic

signal output (standard): 5.0 ml/pulse, 200 pulses/l

max. pump size: 54 / 65 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/497831 / 32822 / 1030

ECOLAB Engineering GmbH
D-83131 Siegedorf



Anlaufgrenze 6 [l/h]*
lower measuring range limit

Impulszahl 35 [n/min]
pulse rate

bei 10 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 29.07.22

Prüfer **P 410**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00540

PN: **200073**

OGM PLUS 00540VCEPPKKE-G1/4I-99-0005m-P

materials of construction: PVC, EPDM, PEEK, ceramic

signal output (standard): 5,0 ml/pulse, 200 pulses/l

max. pump size: 54 / 65 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01 / 497831 / 32822 / 1031

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 6 [l/h]*
lower measuring range limit

Impulszahl 33 [n/min]
pulse rate


bei 10 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 29.07.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00540		
PN: 280073		
OGM PLUS 00540VCEPPKKE-G1/4I-00-0005m-P		
materials of construction:	PVC, EPDM, PEEK, ceramic	
signal output (standard):	5,0 ml/pulse, 200 pulses/l	
max. pump size:	54 / 85 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	01/497831 / 32822 / 1032	
ECOLAB Engineering GmbH D-83313 Siegedorf		

Anlaufgrenze 6 [l/h]*
lower measuring range limit

Impulszahl 32 [n/min]
pulse rate

bei 10 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 29.07.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00540

PN: **280073**

OGM PLUS 00540VCEPPKKE-G1/4I-99-0005m-P

materials of construction: PVC, EPDM, PEEK, ceramic

signal output (standard): 5,0 ml/pulse, 200 pulses/l

max. pump size: 54 / 65 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/497831/32822/1033

ECOLAB Engineering GmbH
D-63313 Sieglar



Anlaufgrenze 6 [l/h]*
lower measuring range limit

Impulszahl 35 [n/min]
pulse rate

bei 10 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 29.07.22

Prüfer P 410