

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 04500		
PN: 280046		
OGM PLUS 04500VCFPPKKE-G5/4I-99-0050m-P		
materials of construction:	PVC, FPM, PEEK, ceramic	
signal output (standard):	50,0 ml/pulse, 20 pulses/l	
max. pump size:	450 / 540 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	02/705592/14122/1001	
ECOLAB Engineering GmbH D-83313 Siegedorf		

Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 41 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 11.10.22

Prüfer P391

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 04500	
PN: 280046	
OGM PLUS 04500VCFPPKKE-G5/4i-99-0050m-P	
materials of construction:	PVC, FPM, PEEK, ceramic
signal output (standard):	50,0 ml/pulse, 20 pulses/l
max. pump size:	450 / 540 l/h (50/60Hz)
semiconductor junction:	PNP
prod. code:	02/705532 /14122 /1002
ECOLAB Engineering GmbH	
D-83313 Siegsdorf	

Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 42 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 11.10.22

Prüfer P39M

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 04500		
PN: 280046		
OGM PLUS 04500VCFPPKKE-G5/4I-99-0050m-P		
materials of construction:	PVC, FPM, PEEK, ceramic	
signal output (standard):	50,0 ml/pulse, 20 pulses/l	
max. pump size:	450 / 540 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	02/705532 /14122 /1008	
ECOLAB Engineering GmbH D-83313 Siegedorf		

Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 42 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 11.10.22

Prüfer P391

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 04500		
PN: 280048		
OGM PLUS 04500VCFPPKKE-G5/41-99-0050m-P		
materials of construction:	PVC, FPM, PEEK, ceramic	
signal output (standard):	50,0 ml/pulse, 20 pulses/l	
max. pump size:	450 / 540 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	02/705532 /14122 /1004	
ECOLAB Engineering GmbH		
D-83313 Siegsdorf		

Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 41 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 11.10.22

Prüfer P391

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 04500
PN: **280048** 
OGM PLUS 04500VCFPPKKE-G5/4I-99-0050m-P
materials of construction: PVC, FPM, PEEK, ceramic
signal output (standard): 50,0 ml/pulse, 20 pulses/l
max. pump size: 450 / 540 l/h (50/60Hz)
semiconductor junction: PNP
prod. code: 02/705532/14+22/1005

ECOLAB Engineering GmbH
D-83313 Siegsdorf 

Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 41 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 11.10.22

Prüfer P391