

# Prüfprotokoll OGM

Oval gear meter type 0.1

PN: **280001**

OGM 001VCFPPKKEI-G1/8i-99-2600-0003

Materials of construction:

PVC, FPM, PEEK,  
ceramic, titan  
2800 pulse/l

Signal output:

Max. flow rate:

Prod. code:

712825 / 51023 / 1001

ECOLAB Engineering GmbH  
D-83313 Slegsdorf



Anlaufgrenze 0,80 [l/h]\*  
lower measuring range limit

Impulszahl 52 [n/min]  
pulse rate

bei 1,25 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 13.3.2023

Prüfer P391

# Prüfprotokoll OGM

Oval gear meter type O.1

PN: **260001**

OGM 001VCFPPKKEI-G1/8i-99-2600-0003

Materials of construction:

PVC, FPM, PEEK,

ceramic, titan

Signal output:

2600 pulse/l

Max. flow rate:

3 l/h

Prod. code:

712625 / 51023/1002

ECOLAB Engineering GmbH

D-83313 Siegsdorf



Anlaufgrenze 0,80 [l/h]\*  
lower measuring range limit

Impulszahl 53 [n/min]  
pulse rate

bei 1,25 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 13.3.2023

Prüfer R301