

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 04500		
PN: 280046		
OGM PLUS 04500VCFPPKKE-G5/4I-99-0050m-P		
materials of construction:	PVC, FPM, PEEK, ceramic	
signal output (standard):	50,0 ml/pulse, 20 pulses/l	
max. pump size:	450 / 540 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	02/497042 /42522 /1001	
ECOLAB Engineering GmbH D-83313 Siegsdorf		

Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 42 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28.06.22

Prüfer **P 410**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 04500

PN: **280048**



OGM PLUS 04500VCFPPKKE-G5/4I-99-0050m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 50,0 ml/pulse, 20 pulses/l

max. pump size: 450 / 540 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 02/497042 /42522 /1002

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 41 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28. 06. 22

Prüfer **P 410**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 04500		
PN: 280046		
OGM PLUS 04500VCFPPKKE-G5/4l-99-0050m-P		
materials of construction:	PVC, FPM, PEEK, ceramic	
signal output (standard):	50,0 ml/pulse, 20 pulses/l	
max. pump size:	450 / 540 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	02/497042 /42522 /1003	
ECOLAB Engineering GmbH		
D-83313 Siegedorf		

Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 39 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28.06.22

Prüfer **P 410**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 04500
PN: **280046** 
OGM PLUS 04500VCFPPKKE-G5/4I-99-0050m-P
materials of construction: PVC, FPM, PEEK, ceramic
signal output (standard): 50,0 ml/pulse, 20 pulses/l
max. pump size: 450 / 540 l/h (50/60Hz)
semiconductor junction: PNP
prod. code: 02/497042/42522/1004

ECOLAB Engineering GmbH
D-83313 Siegsdorf 

Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 41 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28.06.22

Prüfer **P 410**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 04500		
PN: 280046		
OGM PLUS 04500VCFPPKKE-G5/4I-99-0050m-P		
materials of construction:	PVC, FPM, PEEK, ceramic	
signal output (standard):	50,0 ml/pulse, 20 pulses/l	
max. pump size:	450 / 540 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	02/497042 /42522 /1005	
ECOLAB Engineering GmbH D-83313 Siegedorf		

Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 39 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28.06.22

Prüfer **P 410**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 04500

PN: **280048**



OGM PLUS 04500VCFPPKKE-G5/4I-99-0050m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 50,0 ml/pulse, 20 pulses/l

max. pump size: 450 / 540 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 02/497042 / 42522 / 1006

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 40 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28.06.22

Prüfer **P 410**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 04500

PN: **280046**



OGM PLUS 04500VCFPPKKE-G5/4I-99-0050m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 50,0 ml/pulse, 20 pulses/l

max. pump size: 450 / 540 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 02/497042/42522/1007

ECOLAB Engineering GmbH
D-83313 Siegsdorf



Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 43 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28.06.22

Prüfer **P 410**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 04500		
PN: 280046		
OGM PLUS 04500VCFPPKKE-G5/4I-99-0050m-P		
materials of construction:	PVC, FPM, PEEK, ceramic	
signal output (standard):	50,0 ml/pulse, 20 pulses/l	
max. pump size:	450 / 540 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	02/497042/42522/1008	
ECOLAB Engineering GmbH D-83313 Siegedorf		

Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 40 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28.06.22

Prüfer **P 410**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 04500	
PN: 280048	
OGM PLUS 04500VCFPPKKE-G5/4I-99-0050m-P	
materials of construction:	PVC, FPM, PEEK, ceramic
signal output (standard):	50,0 ml/pulse, 20 pulses/l
max. pump size:	450 / 540 l/h (50/60Hz)
semiconductor junction:	PNP
prod. code:	02/497042/42522/1009
ECOLAB Engineering GmbH D-83313 Siegsdorf	

Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 40 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28.06.22

Prüfer **P 410**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 04500

PN: **280048**

OGM PLUS 04500VCFPPKKE-G5/4I-99-0050m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 50,0 ml/pulse, 20 pulses/l

max. pump size: 450 / 540 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 02/497042 /42522 /1010

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 39 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28.06.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 04500

PN: **280046**

OGM PLUS 04500VCFPPKKE-G5/4I-99-0050m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 50,0 ml/pulse, 20 pulses/l

max. pump size: 450 / 540 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 02/497042 /42522 /1011

ECOLAB Engineering GmbH
D-83313 Siegsdorf



Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 41 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28.06.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 04500		
PN: 280048		
OGM PLUS 04500VCFPPKKE-G5/4I-99-0050m-P		
materials of construction:	PVC, FPM, PEEK, ceramic	
signal output (standard):	50,0 ml/pulse, 20 pulses/l	
max. pump size:	450 / 540 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	02/497042 /42522 /1012	
ECOLAB Engineering GmbH D-83313 Siegedorf		

Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 42 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28.06.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 04500		
PN: 280046		
OGM PLUS 04500VCFPPKKE-G5/4I-99-0050m-P		
materials of construction:	PVC, FPM, PEEK, ceramic	
signal output (standard):	50.0 ml/pulse, 20 pulses/l	
max. pump size:	450 / 540 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	02/497042/42522/1013	
ECOLAB Engineering GmbH D-83313 Siegedorf		

Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 41 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28.06.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 04500

PN: **280046**

OGM PLUS 04500VCFPPKKE-G5/4I-99-0050m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 50,0 ml/pulse, 20 pulses/l

max. pump size: 450 / 540 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 02/497042 /42522 /1014

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 42 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28.06.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 04500		
PN: 280048		
OGM PLUS 04500VCFPPKKE-G5/4I-99-0050m-P		
materials of construction:	PVC, FPM, PEEK, ceramic	
signal output (standard):	50,0 ml/pulse, 20 pulses/l	
max. pump size:	450 / 540 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	02/497042/42522/1015	
ECOLAB Engineering GmbH D-83313 Slegsdorf		

Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 42 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28.06.22

Prüfer **P 410**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 04500

PN: **260046**

OGM PLUS 04500VCFPPKKE-G5/4I-99-0050m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 50,0 ml/pulse, 20 pulses/l

max. pump size: 450 / 540 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 02/497042/42522/1016

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 41 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28.06.22

Prüfer **P_410**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 04500		
PN: 280046		
OGM PLUS 04500VCFPPKKE-G5/4I-99-0050m-P		
materials of construction:	PVC, FPM, PEEK, ceramic	
signal output (standard):	50.0 ml/pulse, 20 pulses/l	
max. pump size:	450 / 540 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	02/497042/42522/1017	
ECOLAB Engineering GmbH D-83313 Siegedorf		

Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 43 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28.06.22

Prüfer **P 410**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 04500

PN: **280046**



OGM PLUS 04500VCFPPKKE-G5/4I-99-0050m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 50,0 ml/pulse, 20 pulses/l

max. pump size: 450 / 540 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 02/497042/42522/1018

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 41 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28.06.22

Prüfer **P 410**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 04500	
PN: 280046	
OGM PLUS 04500VCFPPKKE-G5/4I-99-0050m-P	
materials of construction:	PVC, FPM, PEEK, ceramic
signal output (standard):	50,0 ml/pulse, 20 pulses/l
max. pump size:	450 / 540 l/h (50/60Hz)
semiconductor junction:	PNP
prod. code:	02/497042 / 42522 / 1019
ECOLAB Engineering GmbH D-83313 Siegsdorf	

Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 39 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28.06.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 04500

PN: **280046**

OGM PLUS 04500VCFPPKKE-G5/4I-99-0050m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 50,0 ml/pulse, 20 pulses/l

max. pump size: 450 / 540 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 02/497042 /42522 /1020

ECOLAB Engineering GmbH
D-83313 Siegsdorf



Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 42 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28.06.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 04500		
PN: 280048		
OGM PLUS 04500VCFPPKKE-G5/41-99-0050m-P		
materials of construction:	PVC, FPM, PEEK, ceramic	
signal output (standard):	50,0 ml/pulse, 20 pulses/l	
max. pump size:	450 / 540 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	02/497042 /42522 /1021	
ECOLAB Engineering GmbH D-83313 Siegsdorf		

Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 41 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28.06.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 04500

PN: **280046**

OGM PLUS 04500VCFPPKKE-G5/4I-99-0050m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 50,0 ml/pulse, 20 pulses/l

max. pump size: 450 / 540 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 02/497042 / 42522 / 1022

ECOLAB Engineering GmbH

D-83313 Siegedorf



Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 41 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28.06.22

Prüfer **P 410**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 04500		
PN: 280048		
OGM PLUS 04500VCFPPKKE-G5/4I-99-0050m-P		
materials of construction:	PVC, FPM, PEEK, ceramic	
signal output (standard):	50,0 ml/pulse, 20 pulses/l	
max. pump size:	450 / 540 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	02/497042 / 42522 / 1023	
ECCLAB Engineering GmbH D-83313 Siegsdorf		

Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 40 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28.06.22

Prüfer **P 410**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 04500

PN: **280046**

OGM PLUS 04500VCFPPKKE-G5/4I-99-0050m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 50,0 ml/pulse, 20 pulses/l

max. pump size: 450 / 540 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 02/497042 / 42522 / 1024

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 41 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28.06.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 04500		
PN: 280048		
OGM PLUS 04500VCFPPKKE-G5/4l-99-0050m-P		
materials of construction:	PVC, FPM, PEEK, ceramic	
signal output (standard):	50,0 ml/pulse, 20 pulses/l	
max. pump size:	450 / 540 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	02/497042 /42522 /1025	
ECOLAB Engineering GmbH		
D-83313 Siegedorf		

Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 39 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28.06.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 04500		
PN: 280046		
OGM PLUS 04500VCFPPKKE-G5/4l-99-0050m-P		
materials of construction:	PVC, FPM, PEEK, ceramic	
signal output (standard):	50,0 ml/pulse, 20 pulses/l	
max. pump size:	450 / 540 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	02/497042/42522/1026	
ECOLAB Engineering GmbH D-83313 Siegedorf		

Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 41 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28.06.22

Prüfer **P 410**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 04500		
PN: 280046		
OGM PLUS 04500VCFPPKKE-G5/4I-99-0050m-P		
materials of construction:	PVC, FPM, PEEK, ceramic	
signal output (standard):	50,0 ml/pulse, 20 pulses/l	
max. pump size:	450 / 540 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	02/497042 / 42522 / 1027	
ECOLAB Engineering GmbH D-83313 Siegsdorf		

Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 41 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28.06.22

Prüfer **P 410**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 04500

PN: **280046**

OGM PLUS 04500VCFPPKKE-G5/4l-99-0050m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 50,0 ml/pulse, 20 pulses/l

max. pump size: 450 / 540 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 02/497042 /42522 /1028

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 41 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28.06.22

Prüfer **P 410**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 04500

PN: **280046**

OGM PLUS 04500VCFPPKKE-G5/4l-99-0050m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 50.0 ml/pulse, 20 pulses/l

max. pump size: 450 / 540 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 02/497042 /42522 /1029

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 39 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28.06.22

Prüfer **P 410**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 04500		
PN: 280046		
OGM PLUS 04500VCFPPKKE-G5/4I-99-0050m-P		
materials of construction:	PVC, FPM, PEEK, ceramic	
signal output (standard):	50.0 ml/pulse, 20 pulses/l	
max. pump size:	450 / 540 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	02/497042 /42522 /1030	
ECOLAB Engineering GmbH D-83313 Siegedorf		

Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 42 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28.06.22

Prüfer **P 410**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 04500

PN: **280046**



OGM PLUS 04500VCFPPKKE-G5/4I-99-0050m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 50,0 ml/pulse, 20 pulses/l

max. pump size: 450 / 540 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 02/497042 / 42522 / 1031

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 40 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28.06.22

Prüfer **P 410**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 04500

PN: **280048**



OGM PLUS 04500VCFPPKKE-G5/4I-99-0050m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 50,0 ml/pulse, 20 pulses/l

max. pump size: 450 / 540 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 02/497042 / 42522 / 1032

ECOLAB Engineering GmbH
D-83313 Siegsdorf



Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 41 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28.06.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 04500
PN: **280046** 
OGM PLUS 04500VCFPPKKE-G5/41-99-0050m-P
materials of construction: PVC, FPM, PEEK, ceramic
signal output (standard): 50,0 ml/pulse, 20 pulses/l
max. pump size: 450 / 540 l/h (50/60Hz)
semiconductor junction: PNP
prod. code: 02/497042/42522/1033

ECOLAB Engineering GmbH 
D-83313 Siegsdorf

Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 43 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28.06.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 04500	
PN: 280046	
OGM PLUS 04500VCFPPKKE-G5/4I-99-0050m-P	
materials of construction:	PVC, FPM, PEEK, ceramic
signal output (standard):	50,0 ml/pulse, 20 pulses/l
max. pump size:	450 / 540 l/h (50/60Hz)
semiconductor junction:	PNP
prod. code:	02/497042 /42522 /1034
ECOLAB Engineering GmbH	
D-83313 Siedsdorf	

Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 42 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28.06.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 04500		
PN: 280046		
OGM PLUS 04500VCFPPKKE-G5/4I-99-0050m-P		
materials of construction:	PVC, FPM, PEEK, ceramic	
signal output (standard):	50.0 ml/pulse, 20 pulses/l	
max. pump size:	450 / 540 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	02/497042 / 42522 / 1035	
ECLAB Engineering GmbH D-83313 Siegsdorf		

Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 41 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28.06.22

Prüfer **P 410**