


Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 02100		
PN: 280077		
OGM PLUS 02100VCEPPKKE-G5/41-99-0020m-P		
materials of construction:	PVC, EPDM, PEEK, ceramic	
signal output (standard):	20,0 ml/pulse, 50 pulses/l	
max. pump size:	210 / 250 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	02/701833/32922/1001	
ECOLAB Engineering GmbH		
D-83313 Siegedorf		

Anlaufgrenze 50 [l/h]*
lower measuring range limit

Impulszahl 63 [n/min]
pulse rate

bei 80 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 21.07.22

Prüfer P 410