

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200

PN: **280044**



OGM PLUS 01200VCFPPKKE-G3/4I-99-0010m-P

materials of construction: PVC, FPM, PEEK, ceramic
signal output (standard): 10,0 ml/pulse, 100 pulses/l
max. pump size: 120 / 144 l/h (50/60Hz)
semiconductor junction: PNP
prod. code: 01/701849/13522/1001

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl _____ [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 31.08.22

Prüfer P3

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200

PN: **280044**

OGM PLUS 01200VCFPPKKE-G3/4I-99-0010m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 10.0 ml/pulse, 100 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/701849/13522/1002

ECOLAB Engineering GmbH

D-83313 Siegsdorf



Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 55 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 30.08.22

Prüfer P3

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200

PN: 280044

OGM PLUS 01200VCFPPKKE-G3/4I-99-0010m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 10,0 ml/pulse, 100 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01 / 701 849 / 13522 / 1003

ECOLAB Engineering GmbH

D-83313 Siegedorf



Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 55 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 30.08.22

Prüfer P3

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200

PN: **280044**



OGM PLUS 01200VCFPPKKE-G3/4I-99-0010m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 10,0 ml/pulse, 100 pulses/l

max. pump size: 120 / 144 l/h (60/60Hz)

semiconductor junction: PNP

prod. code: 01/701849/13522/1004

ECOLAB Engineering GmbH

D-83313 Blegdorf



Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 57 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 30.08.22

Prüfer P3

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200

PN: 280044

OGM PLUS 01200VCFPPKKE-G3/4I-99-0010m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 10,0 ml/pulse, 100 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/701849/13522/1005

ECOLAB Engineering GmbH

D-83313 Biegendorf



Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 57 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 30.08.22

Prüfer P3

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200

PN: **280044**

OGM PLUS 01200VCFPPKKE-G3/4I-99-0010m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 10,0 ml/pulse, 100 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/701849 / 13522 / 1008

ECOLAB Engineering GmbH
D-83313 Slegsdorf



Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 59 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 30.08.22

Prüfer P3

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200

PN: **280044**



OGM PLUS 01200VCFPPKKE-G3/4I-99-0010m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 10.0 ml/pulse, 100 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/701849/13522/1007

ECOLAB Engineering GmbH
D-83313 Slegsdorf



Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 56 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 30.08.22

Prüfer P3

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200

PN: **280044**



OGM PLUS 01200VCFPPKKE-G3/4I-00-0010m-P

materials of construction: PVC, FPM, PEEK, ceramic
signal output (standard): 10,0 ml/pulse, 100 pulses/l
max. pump size: 120 / 144 l/h (60/60Hz)
semiconductor junction: PNP
prod. code: 01/701849/13522/1008

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 55 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 30.08.22

Prüfer P3

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200

PN: **280044**

OGM PLUS 01200VCFPPKKE-G3/4I-99-0010m-P

materials of construction: PVC, FPM, PEEK, ceramic
signal output (standard): 10,0 ml/pulse, 100 pulses/l
max. pump size: 120 / 144 l/h (50/60Hz)
semiconductor junction: PNP
prod. code: 01/701849/13522/1000

ECOLAB Engineering GmbH
D-83313 Stegedorf



Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 55 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 30.08.22

Prüfer P3

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200

PN: 260044

OGM PLUS 01200VCFPPKKE-G3/4I-99-0010m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 10,0 ml/pulse, 100 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/701849/13522/1010

ECOLAB Engineering GmbH

D-83313 Siegsdorf



Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 56 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 30.08.22

Prüfer P3

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200

PN: **280044**



OGM PLUS 01200VCFPPKKE-G3/4I-99-0010m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 10,0 ml/pulse, 100 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/701849/13522/1011

ECOLAB Engineering GmbH

D-83313 Siegedorf



Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 56 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 30.08.22

Prüfer P3

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200

PN: **280044**



OGM PLUS 01200VCFPPKKE-G3/4I-99-0010m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 10,0 ml/pulse, 100 pulses/l

max. pump size: 120 / 144 l/h (60/60Hz)

semiconductor junction: PNP

prod. code: 01 / 701849 / 13522 / 1012

ECOLAB Engineering GmbH

D-83133 Siegedorf



Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 56 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 30.08.22

Prüfer P3

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200

PN: **280044**



OGM PLUS 01200VCFPPKKE-G3/4I-99-0010m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 10.0 ml/pulse, 100 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/701849/19522/1013

ECOLAB Engineering GmbH

D-83313 Siegedorf



Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 58 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 30.08.22

Prüfer P3

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200

PN: **280044**

OGM PLUS 01200VCFPPKKE-G3/4I-99-0010m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 10.0 ml/pulse, 100 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/701849/13522/1014

ECOLAB Engineering GmbH
D-83313 Biegendorf



Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 55 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 30.08.22

Prüfer P3

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200

PN: **280044**



OGM PLUS 01200VCFPPKKE-G3/4I-99-0010m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 10.0 ml/pulse, 100 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/701849/13522/1015

ECOLAB Engineering GmbH

D-83313 Siegsdorf



Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 56 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 30.08.22

Prüfer P3