

Prüfprotokoll OGM

Oval gear meter type 0.1

PN: 280001

OGM 001VCFPPKKEI-G1/8i-99-2600-0003

Materials of construction:

PVC, FPM, PEEK,
ceramic, titan

Signal output:

2600 pulse/l

Max. flow rate:

3 l/h

Prod. code:

705426 / 14222 / 1004

ECOLAB Engineering GmbH
D-83313 Siedsdorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 55 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 18.10.2022

Prüfer P391

Prüfprotokoll OGM

Oval gear meter type 0.1

PN: 280001

OGM 001VCFPPKKEI-G1/8i-99-2800-0003

Materials of construction:

PVC, FPM, PEEK,
ceramic, titan

Signal output:

2800 pulse/l

Max. flow rate:

3 l/h

Prod. code:

705426 / 142 22 / 1003

ECOLAB Engineering GmbH
D-83313 Siegsdorf



Anlaufgrenze 08 [l/h]*
lower measuring range limit

Impulszahl 54 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 18.10.2022

Prüfer P391

Prüfprotokoll OGM

Oval gear meter type 0.1

PN: 280001

OGM 001VCFPPKKEI-G1/8i-99-2600-0003

Materials of construction:

PVC, FPM, PEEK,
ceramic, titan

Signal output:

2800 pulse/l

Max. flow rate:

3 l/h

Prod. code:

705426 / 14222 / 1002

ECOLAB Engineering GmbH
D-83313 Siegsdorf



Anlaufgrenze 0.8 [l/h]*
lower measuring range limit

Impulszahl 52 [n/min]
pulse rate

bei 1.25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 18.10.2022

Prüfer P391

Prüfprotokoll OGM

Oval gear meter type 0.1

PN: 280001

OGM 001VCFFPKKETI-G1/8i-99-2600-0003

Materials of construction:

PVC, FPM, PEEK,
ceramic, titan

Signal output:

2800 pulse/l

Max. flow rate:

3 l/h

Prod. code:

705426 / 14222 / 1001

ECOLAB Engineering GmbH
D-83313 Siegsdorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 51 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 18.10.2022

Prüfer P391