

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 01200

PN: **280048**

OGM PLUS 01200PPEPPKKE-G1/2I-99-0005m-P

materials of construction: PP, EDPM, PEEK, ceramic

signal output (standard): 5,0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 493208 / 21222 / 1030

ECOLAB Engineering GmbH  
D-83313 Siegedorf



Anlaufgrenze 20 [l/h]\*  
lower measuring range limit

Impulszahl 124 [n/min]  
pulse rate

bei 35 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 30.3.2022

Prüfer P391

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 01200

PN: **280048**

OGM PLUS 01200PPEPPKKE-G1/2I-99-0005m-P

materials of construction: PP, EDPM, PEEK, ceramic

signal output (standard): 5,0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 493209/21222/1029

ECOLAB Engineering GmbH  
D-83313 Siegedorf



Anlaufgrenze 20 [l/h]\*  
lower measuring range limit

Impulszahl 115 [n/min]  
pulse rate

bei 35 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 30.3.2022

Prüfer P391

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 01200

PN: 280048

OGM PLUS 01200PPEPPKKE-G1/2I-99-0005m-P

materials of construction: PP, EDPM, PEEK, ceramic

signal output (standard): 5.0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 493206 / 21222 / 1028

ECOLAB Engineering GmbH  
D-83313 Stegedorf



Anlaufgrenze 20 [l/h]\*  
lower measuring range limit

Impulszahl 121 [n/min]  
pulse rate

bei 35 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 30.3.2022

Prüfer P391

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 01200



PN: **280048**

OGM PLUS 01200PPEPPKKE-G1/2I-99-0005m-P

materials of construction:	PP, EDPM, PEEK, ceramic
signal output (standard):	5,0 ml/pulse, 200 pulses/l
max. pump size:	120 / 144 l/h (50/60Hz)
semiconductor junction:	PNP
prod. code:	493206 / 21222 / 1027

ECOLAB Engineering GmbH  
D-83313 Siegedorf



Anlaufgrenze 20 [l/h]\*  
lower measuring range limit

Impulszahl 123 [n/min]  
pulse rate

bei 35 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 30.3.2022

Prüfer R391

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 01200

PN: **280048**

OGM PLUS 01200PPEPPKKE-G1/2I-99-0005m-P

materials of construction: PP, EDPM, PEEK, ceramic

signal output (standard): 5,0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 463206/21222/1026

ECOLAB Engineering GmbH  
D-83313 Siegsdorf



Anlaufgrenze 20 [l/h]\*  
lower measuring range limit

Impulszahl 120 [n/min]  
pulse rate

bei 35 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 30.3.2022

Prüfer P391

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 01200

PN: **280048**

OGM PLUS 01200PPEPPKKE-G1/2I-99-0005m-P

materials of construction: PP, EDPM, PEEK, ceramic

signal output (standard): 5.0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 493206 / 21 222 / 1025

ECOLAB Engineering GmbH  
D-63313 Siegedorf



Anlaufgrenze 20 [l/h]\*  
lower measuring range limit

Impulszahl 121 [n/min]  
pulse rate

bei 35 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 30.3.2022

Prüfer P391

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 01200

PN: **280048**



OGM PLUS 01200PPEPPKKE-G1/2I-99-0005m-P

materials of construction: PP, EDPM, PEEK, ceramic

signal output (standard): 5,0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: **493206 / 21222 / 1024**

ECOLAB Engineering GmbH  
D-83313 Siegedorf



Anlaufgrenze 20 [l/h]\*  
lower measuring range limit

Impulszahl 123 [n/min]  
pulse rate

bei 35 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 30.3.2022

Prüfer P391

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 01200

PN: 280048

OGM PLUS 01200PPEPPKKE-G1/2I-99-0005m-P

materials of construction: PP, EDPM, PEEK, ceramic

signal output (standard): 5.0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 493206 / 21 222 / 1023

ECOLAB Engineering GmbH  
D-83313 Siegedorf



Anlaufgrenze 20 [l/h]\*  
lower measuring range limit

Impulszahl 120 [n/min]  
pulse rate

bei 35 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 30.3.2022

Prüfer P391



# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 01200

PN: **280048**



OGM PLUS 01200PPEPPKKE-G1/2I-99-0005m-P

materials of construction: PP, EDPM, PEEK, ceramic

signal output (standard): 5.0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 493206 / 21222 / 1022

ECOLAB Engineering GmbH  
D-83313 Siegedorf



Anlaufgrenze 20 [l/h]\*  
lower measuring range limit

Impulszahl 119 [n/min]  
pulse rate

bei 35 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 30.3.2022

Prüfer P391

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 01200

PN: **280048**



OGM PLUS 01200PPEPPKKE-G1/2I-99-0005m-P

materials of construction: PP, EDPM, PEEK, ceramic

signal output (standard): 5.0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 493208/21222/1021

ECOLAB Engineering GmbH  
D-63313 Sieglar



Anlaufgrenze 20 [l/h]\*  
lower measuring range limit

Impulszahl 123 [n/min]  
pulse rate

bei 35 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 30.3.2022

Prüfer P391

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 01200

PN: **280048**

OGM PLUS 01200PPEPPKKE-G1/2I-99-0005m-P

materials of construction: PP, EDPM, PEEK, ceramic

signal output (standard): 5,0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 493206 / 21222 / 1020

ECOLAB Engineering GmbH  
D-83313 Siegsdorf

Anlaufgrenze 20 [l/h]\*  
lower measuring range limit

Impulszahl 119 [n/min]  
pulse rate

bei 35 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 30.3.2022

Prüfer P391

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 01200

PN: **280048**

OGM PLUS 01200PPEPPKKE-G1/2I-99-0005m-P

materials of construction: PP, EDPM, PEEK, ceramic

signal output (standard): 5,0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 493206 / 21222 / 1019

ECOLAB Engineering GmbH  
D-83313 Siegedorf



Anlaufgrenze 20 [l/h]\*  
lower measuring range limit

Impulszahl 123 [n/min]  
pulse rate

bei 35 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 30.3.2022

Prüfer P391

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 01200

PN: **280048**

OGM PLUS 01200PPEPPKKE-G1/2I-99-0005m-P

materials of construction: PP, EDPM, PEEK, ceramic

signal output (standard): 5.0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 493206 / 21 222 / 101 B

ECOLAB Engineering GmbH  
D-83313 Siegsdorf



Anlaufgrenze 20 [l/h]\*  
lower measuring range limit

Impulszahl 124 [n/min]  
pulse rate

bei 35 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 30.3.2022

Prüfer P391

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 01200

PN: **280048**

OGM PLUS 01200PPEPPKKE-G1/2I-99-0005m-P

materials of construction: PP, EDPM, PEEK, ceramic

signal output (standard): 5,0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: **493206/21222/1017**

ECOLAB Engineering GmbH  
D-83313 Siegedorf



Anlaufgrenze 20 [l/h]\*  
lower measuring range limit

Impulszahl 126 [n/min]  
pulse rate

bei 35 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 30.3.2022

Prüfer P391

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 01200

PN: **280048**



OGM PLUS 01200PPEPPKKE-G1/2I-99-0005m-P

materials of construction: PP, EDPM, PEEK, ceramic

signal output (standard): 5.0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 493205/21222/1016

ECOLAB Engineering GmbH  
D-83313 Slegsdorf



Anlaufgrenze 20 [l/h]\*  
lower measuring range limit

Impulszahl 124 [n/min]  
pulse rate

bei 35 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 30.3.2022

Prüfer P391

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 01200

PN: **280048**

OGM PLUS 01200PPEPPKKE-G1/2l-99-0005m-P

materials of construction: PP, EDPM, PEEK, ceramic

signal output (standard): 5,0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 493206 / 21222 / 1015

ECOLAB Engineering GmbH  
D-83313 Siegedorf



Anlaufgrenze 20 [l/h]\*  
lower measuring range limit

Impulszahl 114 [n/min]  
pulse rate

bei 35 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 30.3.2022

Prüfer P391



# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 01200

PN: **280048**

OGM PLUS 01200PPEPPKKE-G1/2I-99-0005m-P

materials of construction: PP, EDPM, PEEK, ceramic

signal output (standard): 5,0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 493208/21222/1014

ECOLAB Engineering GmbH  
D-83313 Siegsdorf



Anlaufgrenze 20 [l/h]\*  
lower measuring range limit

Impulszahl 121 [n/min]  
pulse rate

bei 35 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 30.3.2022

Prüfer P391

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 01200

PN: **280048**

OGM PLUS 01200PPEPPKKE-G1/2I-99-0005m-P

materials of construction: PP, EDPM, PEEK, ceramic

signal output (standard): 5,0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 493206 / 21222 / 1013

ECOLAB Engineering GmbH  
D-83313 Siegedorf

Anlaufgrenze 20 [l/h]\*  
lower measuring range limit

Impulszahl 123 [n/min]  
pulse rate

bei 35 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 30.3.2022

Prüfer P391

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 01200

PN: **280048**



OGM PLUS 01200PPEPPKKE-G1/2I-99-0005m-P

materials of construction: PP, EDPM, PEEK, ceramic

signal output (standard): 5,0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 493208 / 21222 / 1012

ECOLAB Engineering GmbH

D-83313 Slegsdorf



Anlaufgrenze 20 [l/h]\*  
lower measuring range limit

Impulszahl 124 [n/min]  
pulse rate

bei 35 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 30.3.2022

Prüfer P391

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 01200

PN: 280048

OGM PLUS 01200PPEPPKKE-G1/2I-00-0005m-P

materials of construction: PP, EDPM, PEEK, ceramic

signal output (standard): 5,0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 493206/21222/1011

ECOLAB Engineering GmbH  
D-83313 Siegedorf



Anlaufgrenze 20 [l/h]\*  
lower measuring range limit

Impulszahl 122 [n/min]  
pulse rate

bei 35 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 29.3.2022

Prüfer P391

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 01200



PN: 280048

OGM PLUS 01200PPEPPKKE-G1/2i-99-0005m-P

materials of construction: PP, EDPM, PEEK, ceramic

signal output (standard): 5.0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 493206/21222/1010

ECOLAB Engineering GmbH  
D-83313 Siegedorf



Anlaufgrenze 20 [l/h]\*  
lower measuring range limit

Impulszahl 124 [n/min]  
pulse rate

bei 35 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 29.3.2022

Prüfer P391

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 01200

PN: **280048**

OGM PLUS 01200PPEPPKKE-G1/2I-99-0005m-P

materials of construction: PP, EDPM, PEEK, ceramic

signal output (standard): 5,0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: **493208/21222/1009**



ECOLAB Engineering GmbH  
D-83313 Siegedorf



Anlaufgrenze 20 [l/h]\*  
lower measuring range limit

Impulszahl 122 [n/min]  
pulse rate

bei 35 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 29.3.2022

Prüfer P391

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 01200

PN: **280048**

OGM PLUS 01200PPEPPKKE-G1/2I-99-0005m-P

materials of construction: PP, EDPM, PEEK, ceramic

signal output (standard): 5,0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 493208 / 21222 / 1008

ECOLAB Engineering GmbH  
D-83313 Siegedorf



Anlaufgrenze 20 [l/h]\*  
lower measuring range limit

Impulszahl 120 [n/min]  
pulse rate

bei 35 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 29.3.2022

Prüfer P391

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 01200

PN: **280048**

OGM PLUS 01200PPEPPKKE-G1/2I-99-0005m-P

materials of construction: PP, EDPM, PEEK, ceramic

signal output (standard): 5.0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (60/60Hz)

semiconductor junction: PNP

prod. code: 493206 / 21 222 / 1007

ECOLAB Engineering GmbH  
D-83313 Siegedorf



Anlaufgrenze 20 [l/h]\*  
lower measuring range limit

Impulszahl 119 [n/min]  
pulse rate

bei 35 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 29.3.2022

Prüfer P391



# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 01200

PN: **280048**

OGM PLUS 01200PPEPPKKE-G1/2I-99-0005m-P

materials of construction: PP, EDPM, PEEK, ceramic

signal output (standard): 5,0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 493205/21222/1006



ECOLAB Engineering GmbH  
D-83313 Siegsdorf



Anlaufgrenze 20 [l/h]\*  
lower measuring range limit

Impulszahl 122 [n/min]  
pulse rate

bei 35 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 29.3.2022

Prüfer P391

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 01200

PN: **280048**



OGM PLUS 01200PPEPPKKE-G1/2I-99-0005m-P

materials of construction: PP, EDPM, PEEK, ceramic

signal output (standard): 5,0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 493206 / 21222 / 1005

ECOLAB Engineering GmbH  
D-83313 Siegsdorf



Anlaufgrenze 20 [l/h]\*  
lower measuring range limit

Impulszahl 115 [n/min]  
pulse rate

bei 35 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 29.3.2022

Prüfer P391

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 01200

PN: **280048**

OGM PLUS 01200PPEPPKKE-G1/2I-99-0005m-P

materials of construction: PP, EDPM, PEEK, ceramic

signal output (standard): 5,0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 493208/21222/1004

ECOLAB Engineering GmbH

D-83313 Siegsdorf



Anlaufgrenze 20 [l/h]\*  
lower measuring range limit

Impulszahl 123 [n/min]  
pulse rate

bei 35 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 29.3.2022

Prüfer P391

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 01200

PN: **280048**

OGM PLUS 01200PPEPPKKE-G1/2I-99-0005m-P

materials of construction: PP, EDPM, PEEK, ceramic

signal output (standard): 5,0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 493206 / 21222 / 1003

ECOLAB Engineering GmbH  
D-83313 Slegsdorf



Anlaufgrenze 20 [l/h]\*  
lower measuring range limit

Impulszahl 124 [n/min]  
pulse rate

bei 35 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 29.3.2022

Prüfer P391

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 01200

PN: **280048**

OGM PLUS 01200PPEPPKKE-G1/2I-99-0005m-P

materials of construction: PP, EDPM, PEEK, ceramic

signal output (standard): 5.0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 493206/21222/1002

ECOLAB Engineering GmbH  
D-83313 Siegedorf



Anlaufgrenze 20 [l/h]\*  
lower measuring range limit

Impulszahl 120 [n/min]  
pulse rate

bei 35 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 29.3.2022

Prüfer P391

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 01200

PN: **280048**



OGM PLUS 01200PPEPPKKE-G1/2I-99-0005m-P

materials of construction: PP, EDPM, PEEK, ceramic

signal output (standard): 5.0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 493206 / 21222 / 1001

ECOLAB Engineering GmbH  
D-63313 Sieddorf



Anlaufgrenze 20 [l/h]\*  
lower measuring range limit

Impulszahl 118 [n/min]  
pulse rate

bei 35 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 29.3.2022

Prüfer P391