

# Prüfprotokoll OGM

Washtronic EXACT II Typ1  
PN: 280032  
OGM 010VCFPPKKEI-G1/4I-99-0820-0030  
Materials of construction: PVC, FPM, PEEK, ceramic, titan  
Signal output: 820 pulse/l  
Max. flow rate: 30 l/h  
Prod. code: 704520 / 241 22 / 1001  
ECOLAB Engineering GmbH  
D-83313 Slegsdorf

Anlaufgrenze 5,5 [l/h]\*  
lower measuring range limit

Impulszahl 119 [n/min]  
pulse rate

bei 9,0 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 12.10.22

Prüfer P391

# Prüfprotokoll OGM

Washtronic EXACT II Typ1

PN: **280032**

OGM 010VCFPPKKEI-G1/4I-99-0820-0030

Materials of construction:

PVC, FPM, PEEK,  
ceramic, titan

Signal output:

820 pulse/l

Max. flow rate:

30 l/h

Prod. code:

704520 / 241 22 / 1002

ECOLAB Engineering GmbH

D-83313 Siegedorf



Anlaufgrenze 5,5 [l/h]\*  
lower measuring range limit

Impulszahl 122 [n/min]  
pulse rate

bei 9,0 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 12.10.22

Prüfer P391