


Prüfprotokoll OGM

Oval gear meter type 2
PN: 280003
OGM 020VCFPPKKEI-G1/4i-99-0450-0060
Materials of construction: PVC, FPM, PEEK, ceramic, titan
Signal output: 450 pulse/l
Max. flow rate: 60 l/h
Prod. code: 495712 / 21522 / 1001
ECOLAB Engineering GmbH
D-83313 Siegsdorf



Anlaufgrenze 11,0 [l/h]*
lower measuring range limit

Impulszahl 132 [n/min]
pulse rate

bei 18,0 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 19.4.22

Prüfer P12

Prüfprotokoll OGM

Oval gear meter type 2

PN: **280003**

OGM 020VCFPPKKEI-G1/4i-99-0450-0060

Materials of construction:

PVC, FPM, PEEK,
ceramic, titan
450 pulse/l
60 l/h

Signal output:

Max. flow rate:

Prod. code:

495712 / 21522 / 1002



ECOLAB Engineering GmbH
D-83313 Siegsdorf

Anlaufgrenze 11,0 [l/h]*
lower measuring range limit

Impulszahl 136 [n/min]
pulse rate

bei 18,0 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 19.4.22

Prüfer P12

Prüfprotokoll OGM

Oval gear meter type 2

PN: 280003

OGM 020VCFPPKKEI-G1/4i-99-0450-0060

Materials of construction:

PVC, FPM, PEEK,
ceramic, titan

Signal output:

450 pulse/l

Max. flow rate:

80 l/h

Prod. code:

495712 / 21522 / 1003

ECOLAB Engineering GmbH
D-83313 Siedsdorf



Anlaufgrenze 11.0 [l/h]*
lower measuring range limit

Impulszahl 130 [n/min]
pulse rate

bei 18.0 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 19.4.22

Prüfer P12