

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 04500		
PN: 290078		
OGM PLUS 04500VCEPPKKE-G5/4i-99-0050m-P		
materials of construction:	PVC, EPDM, PEEK, ceramic	
signal output (standard):	50,0 ml/pulse, 20 pulses/l	
max. pump size:	450 / 540 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	02/497040/32522/1001	
ECOLAB Engineering GmbH D-83313 Siegedorf		

Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 42 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 24.06.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 04500

PN: **280078**

OGM PLUS 04500VCEPPKKE-G5/4I-99-0050m-P

materials of construction: PVC, EPDM, PEEK, ceramic

signal output (standard): 50,0 ml/pulse, 20 pulses/l

max. pump size: 450 / 540 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 02/497040/32522/1002

ECOLAB Engineering GmbH
D-63313 Siegborf



Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 41 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 26.06.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 04500		
PN: 280078		
OGM PLUS 04500VCEPPKKE-G5/4I-99-0050m-P		
materials of construction:	PVC, EPDM, PEEK, ceramic	
signal output (standard):	50,0 ml/pulse, 20 pulses/l	
max. pump size:	450 / 540 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	02/497040/32522/1003	
ECOLAB Engineering GmbH		
D-83313 Siedsdorf		

Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 40 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 24.06.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 04500

PN: **280078**

OGM PLUS 04500VCEPPKKE-G5/4l-99-0050m-P

materials of construction: PVC, EPDM, PEEK, ceramic

signal output (standard): 50,0 ml/pulse, 20 pulses/l

max. pump size: 450 / 540 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 02/497040 / 32522 / 1004

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 39 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 24.06.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 04500

PN: **280078**

OGM PLUS 04500VCEPPKKE-G5/4I-99-0050m-P

materials of construction: PVC, EPDM, PEEK, ceramic

signal output (standard): 50,0 ml/pulse, 20 pulses/l

max. pump size: 450 / 540 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 02/497040/32522/1005

ECOLAB Engineering GmbH
D-63313 Siegsdorf



Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 40 [n/min]
pulse rate


bei 170 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 26.06.22

Prüfer **P 410**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 04500		
PN: 280078		
OGM PLUS 04500VCEPPKKE-G5/4I-99-0050m-P		
materials of construction:	PVC, EPDM, PEEK, ceramic	
signal output (standard):	50.0 ml/pulse, 20 pulses/l	
max. pump size:	450 / 540 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	02/497040/32522/1008	
ECOLAB Engineering GmbH D-83313 Siegedorf		

Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 41 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 24.06.22

Prüfer **P 410**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 04500

PN: **280078**



OGM PLUS 04500VCEPPKKE-G5/4I-99-0050m-P

materials of construction: PVC, EPDM, PEEK, ceramic

signal output (standard): 50.0 ml/pulse, 20 pulses/l

max. pump size: 450 / 540 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 02/497040/32522/1007

ECOLAB Engineering GmbH
D-63313 Siegedorf



Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 40 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 24.06.22

Prüfer **P 410**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 04500

PN: **280078**

OGM PLUS 04500VCEPPKKE-G5/4I-99-0050m-P

materials of construction: PVC, EPDM, PEEK, ceramic

signal output (standard): 50.0 ml/pulse, 20 pulses/l

max. pump size: 450 / 540 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 02/497040/32522/1008

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 39 [n/min]
pulse rate


bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 24.06.22

Prüfer **P 410**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 04500		
PN: 280078		
OGM PLUS 04500VCEPPKKE-G5/4I-99-0050m-P		
materials of construction:	PVC, EPDM, PEEK, ceramic	
signal output (standard):	50,0 ml/pulse, 20 pulses/l	
max. pump size:	450 / 540 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	02/497040/32522/1009	
ECOLAB Engineering GmbH D-83313 Siegedorf		

Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 40 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 24.06.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 04500

PN: **280078**



OGM PLUS 04500VCEPPKKE-G5/4I-99-0050m-P

materials of construction: PVC, EPDM, PEEK, ceramic

signal output (standard): 50.0 ml/pulse, 20 pulses/l

max. pump size: 450 / 540 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 02/497040/32522/1010

ECOLAB Engineering GmbH
D-83313 Siegsdorf



Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 39 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 24.06.22

Prüfer **P 410**