

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**

OGM PLUS 00112VCEPVCKE-G1/8i-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/497241/12722/1001

ECOLAB Engineering GmbH

D-83313 Siegsdorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 22 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 08.07.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**



OGM PLUS 00112VCEPVOKE-G1/81-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/497241 / 12722 / 1002

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 19 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 08.07.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **200151**

OGM PLUS 00112VCEPVCKE-G1/8I-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/497241 / 12722 / 1003

ECOLAB Engineering GmbH
D-83313 Slegsdorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 20 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 08.07.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**



OGM PLUS 00112VCEPVCKE-G1/8I-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/497241/12722/1004

ECOLAB Engineering GmbH
D-83313 Stegsdorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 22 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 08.07.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**



OGM PLUS 00112VCEPVCKE-G1/8I-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/497241 / 12722 / 1005

ECOLAB Engineering GmbH
D-83313 Siegsdorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 19 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 08.07.22

Prüfer **P 410**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**

OGM PLUS 00112VCEPVCKE-G1/8I-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/497241 / 12722 / 1006

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 20 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 08.07.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**

OGM PLUS 00112VCEPVCKE-G1/8I-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/497241 / 12722 / 1007

ECOLAB Engineering GmbH
D-83313 Siegsdorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 13 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 08.07.22

Prüfer **P 410**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **290151**

OGM PLUS 00112VCEPVCKE-G1/BI-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1.0 ml/pulse, 1000 pulses/l

max. pump size: 11.2 / 13.5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/497241/12722/1008

ECOLAB Engineering GmbH
D-83133 Siegsdorf



Anlaufgrenze 0.8 [l/h]*
lower measuring range limit

Impulszahl 21 [n/min]
pulse rate

bei 1.25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 08.07.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**

OGM PLUS 00112VCEPVCKE-G1/8I-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/497241 / 12722 / 1009

ECOLAB Engineering GmbH
D-83313 Stegedorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 19 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 08.07.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**



OGM PLUS 00112VCEPVCKE-G1/8I-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/497241/12722/1010

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 21 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 08.07.22

Prüfer **P 410**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**



OGM PLUS 00112VCEPVCKE-G1/8I-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/497241 / 12722 / 1011

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 21 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 08.07.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**



OGM PLUS 00112VCEPVCKE-G1/8I-99-0001 m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/497241 /12722 /1012

ECOLAB Engineering GmbH
D-83313 Siegsdorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 19 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 08.07.22

Prüfer **P 410**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**



OGM PLUS 00112VCEPVCKE-G1/8I-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/497241 / 12722 / 1013

ECOLAB Engineering GmbH
D-83313 Siegsdorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 21 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 08.07.22

Prüfer **P 410**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**

OGM PLUS 00112VCEPVCKE-G1/8I-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/497241 / 12722 / 1014

ECOLAB Engineering GmbH
D-83313 Siegsdorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 20 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 08.07.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**

OGM PLUS 00112VCEPVCKE-G1/8I-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/497241 / 12722 / 1015

ECOLAB Engineering GmbH
D-83313 Siedsdorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 22 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 08.07.22

Prüfer **P 410**