

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 02100		
PN: 280065		
OGM PLUS 02100VCFPPKKE-G5/4I-99-0125p-P		
materials of construction:	PVC, FPM, PEEK, ceramic	
signal output (standard):	8,0 ml/pulse, 125 pulses/l	
max. pump size:	210 / 250 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	700431 /32922 /1002	
ECOLAB Engineering GmbH D-83313 Siegedorf		

Anlaufgrenze 50 [l/h]*
lower measuring range limit

Impulszahl 188 [n/min]
pulse rate

bei 80 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 21.07.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 02100

PN: **280065**

OGM PLUS 02100VCFPPKKE-G5/4I-99-0125p-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 8.0 ml/pulse, 125 pulses/l

max. pump size: 210 / 250 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 700431 / ~~32622~~ / 1001

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 50 [l/h]*
lower measuring range limit

Impulszahl ~~186~~ 186 [n/min]
pulse rate

bei 80 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 21.07.22

Prüfer **P 410**