

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 04500

PN: **280078**

OGM PLUS 04500VCEPPKKE-G5/4I-99-0050m-P

materials of construction: PVC, EPDM, PEEK, ceramic

signal output (standard): 50.0 ml/pulse, 20 pulses/l

max. pump size: 450 / 540 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 02/704945/13822/1001

ECOLAB Engineering GmbH  
D-83313 Siegedorf



Anlaufgrenze 75 [l/h]\*  
lower measuring range limit

Impulszahl 40 [n/min]  
pulse rate


bei 120 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 22.09.22

Prüfer P 410

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 04500		
PN: <b>280078</b>		
OGM PLUS 04500VCEPPKKE-G5/4I-99-0050m-P		
materials of construction:	PVC, EPDM, PEEK, ceramic	
signal output (standard):	50,0 ml/pulse, 20 pulses/l	
max. pump size:	450 / 540 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	02/704945/13822/1002	
ECOLAB Engineering GmbH D-83313 Siegsdorf		

Anlaufgrenze 75 [l/h]\*  
lower measuring range limit

Impulszahl 42 [n/min]  
pulse rate



bei 120 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 22.09.22

Prüfer P 410

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 04500		
PN: <b>280078</b>		
OGM PLUS 04500VCEPPKKE-G5/4I-99-0050m-P		
materials of construction:	PVC, EPDM, PEEK, ceramic	
signal output (standard):	50.0 ml/pulse, 20 pulses/l	
max. pump size:	450 / 540 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	02/704945/13822/1003	
ECOLAB Engineering GmbH D-63313 Siegedorf		

Anlaufgrenze 75 [l/h]\*  
lower measuring range limit

Impulszahl 42 [n/min]  
pulse rate



bei 120 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 22.09.22

Prüfer P 410

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 04500		
PN: <b>280078</b>		
OGM PLUS 04500VCEPPKKE-G5/4I-99-0050m-P		
materials of construction:	PVC, EPDM, PEEK, ceramic	
signal output (standard):	50.0 ml/pulse, 20 pulses/l	
max. pump size:	450 / 540 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	02/704945/13822/1004	
ECOLAB Engineering GmbH D-83313 Siegedorf		

Anlaufgrenze 75 [l/h]\*  
lower measuring range limit

Impulszahl 41 [n/min]  
pulse rate

bei 120 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 22.05.22

Prüfer P 410

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 04500

PN: **280078**

OGM PLUS 04500VCEPPKKE-G5/4I-99-0050m-P

materials of construction: PVC, EPDM, PEEK, ceramic

signal output (standard): 50.0 ml/pulse, 20 pulses/l

max. pump size: 450 / 540 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 02/704945/13822/1005

ECOLAB Engineering GmbH  
D-83313 Siegsdorf



Anlaufgrenze 75 [l/h]\*  
lower measuring range limit

Impulszahl 41 [n/min]  
pulse rate

bei 120 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 22.05.22

Prüfer P 410