

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**

OGM PLUS 00112VCEPVCKE-G1/8I-99-0001m-P

| | |
|----------------------------|-----------------------------|
| materials of construction: | PVC, EPDM, PVDF, ceramic |
| signal output (standard): | 1,0 ml/pulse, 1000 pulses/l |
| max. pump size: | 11,2 / 13,5 l/h (50/60Hz) |
| semiconductor junction: | PNP |
| prod. code: | 01/706011/54822/1001 |

ECOLAB Engineering GmbH
D-93313 Siegsdorf

Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 19 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 12.12.22

Prüfer Sollereder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**

OGM PLUS 00112VCEPVCKE-G1/8I-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/706011/54822/1002

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 19 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 12.12.22

Prüfer Sollerseder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: 280151



OGM PLUS 00112VCEPVCKE-G1/81-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic
signal output (standard): 1,0 ml/pulse, 1000 pulses/l
max. pump size: 11,2 / 13,5 l/h (50/60Hz)
semiconductor junction: PNP
prod. code: 01/706011/54822/1003

ECOLAB Engineering GmbH
D-83313 Siegsdorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 19 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 12.12.22

Prüfer Solleroder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**

OGM PLUS 00112VCEPVCKE-G1/8I-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/708011 / 54822 / 1004

ECOLAB Engineering GmbH
D-83313 Slegsdorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 19 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 13.12.22

Prüfer Sollereder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**

OGM PLUS 00112VCEPVCKE-G1/8I-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/706011 / 54822 / 1005

ECOLAB Engineering GmbH
D-63313 Siegedorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 19 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 13.12.22

Prüfer Sollweider

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**

OGM PLUS 00112VCEPVCKE-G1/8I-09-0001m-P

| | |
|----------------------------|-----------------------------|
| materials of construction: | PVC, EPDM, PVDF, ceramic |
| signal output (standard): | 1,0 ml/pulse, 1000 pulses/l |
| max. pump size: | 11,2 / 19,5 l/h (50/60Hz) |
| semiconductor junction: | PNP |
| prod. code: | 01/706011 / 54822 / 1006 |

ECOLAB Engineering GmbH
D-83313 Siegsdorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 19 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 13.12.22

Prüfer Sollerseder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**

OGM PLUS 00112VCEPVCKE-G1/8I-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/708011 / 54822 / 1007

ECOLAB Engineering GmbH
D-63313 Siegedorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 19 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 13.12.22

Prüfer _____

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**

OGM PLUS 00112VCEPVCKE-G1/8I-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/708011/54822/1008

ECOLAB Engineering GmbH
D-83313 Siegsdorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 19 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 13.12.22

Prüfer Sollereder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**

OGM PLUS 00112VCEPVCKE-G1/81-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/708011 / 54822 / 1009

ECOLAB Engineering GmbH
D-83313 Siegsdorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 20 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 13.12.22

Prüfer Sollender

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**



OGM PLUS 00112VCEPVCKE-G1/8l-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: Q1/706011 / 54822 / 1010

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 79 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 13.12.22

Prüfer Sollerseder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**

OGM PLUS 00112VCEPVCKE-G1/8I-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic
signal output (standard): 1,0 ml/pulse, 1000 pulses/l
max. pump size: 11,2 / 13,5 l/h (50/60Hz)
semiconductor junction: PNP
prod. code: 01/706011 / 54822 / 1011

ECOLAB Engineering GmbH
D-83313 Slegsdorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 19 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 13.12.22

Prüfer Sollerseder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**



OGM PLUS 00112VCEPVCKE-G1/8I-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/708011 / 54822 / 1012

ECOLAB Engineering GmbH
D-83313 Slegendorf



Anlaufgrenze 9,8 [l/h]*
lower measuring range limit

Impulszahl 19 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 13.12.22

Prüfer Sollereder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **290151**

OGM PLUS 00112VCEPVCKE-G1/8i-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/709011 / 54822 / 1013

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 19 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 13.12.22

Prüfer Sollereder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**

OGM PLUS 00112VCEPVCKE-G1/8I-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic
signal output (standard): 1,0 ml/pulse, 1000 pulses/l
max. pump size: 11,2 / 13,5 l/h (50/60Hz)
semiconductor junction: PNP
prod. code: 01/708011 / 54822 / 1014

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 19 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 13.12.22

Prüfer Sollereder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112



PN: **280151**

OGM PLUS 00112VCEPVCKE-G1/8I-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/708011/54822/1015

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 20 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 13.12.22

Prüfer Johannes

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**

OGM PLUS 00112VCEPVCKE-G1/8I-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/706011/54822/1018

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 9,8 [l/h]*
lower measuring range limit

Impulszahl 20 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 13.12.22

Prüfer Sollereder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**

OGM PLUS 00112VCEPVCKE-G1/8i-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/708011/54822/1017

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 20 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 13.12.22

Prüfer Jollereder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**

OGM PLUS 00112VCEPVCKE-G1/8I-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/708011 / 54822 / 1018

ECOLAB Engineering GmbH
D-93313 Siegsdorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 19 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 13.12.22

Prüfer Sollerseder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**

OGM PLUS 00112VCEPVCKE-G1/8I-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/706011/54822/1019

ECOLAB Engineering GmbH
D-83313 Siegsdorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 19 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 13.12.22

Prüfer Sollereder