

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 02100

PN: **280045**

OGM PLUS 02100VCFPPKKE-G5/4I-99-0020m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 20,0 ml/pulse, 50 pulses/l

max. pump size: 210 / 250 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 02/701863/33422/1001

ECOLAB Engineering GmbH  
D-83313 Siegedorf



Anlaufgrenze 50 [l/h]\*  
lower measuring range limit

Impulszahl 70 [n/min]  
pulse rate

bei 80 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 30.08.22

Prüfer P3

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 02100

PN: **280045**

OGM PLUS 02100VCFPPKKE-G5/4I-99-0020m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 20.0 ml/pulse, 50 pulses/l

max. pump size: 210 / 250 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 02/701893/33422/1002

ECOLAB Engineering GmbH  
D-83313 Siegsdorf



Anlaufgrenze 50 [l/h]\*  
lower measuring range limit

Impulszahl 73 [n/min]  
pulse rate

bei 80 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 30.08.22

Prüfer P3

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 02100

PN: 280045



OGM PLUS 02100VCFPPKKE-G5/4I-99-0020m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 20,0 ml/pulse, 50 pulses/l

max. pump size: 210 / 250 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 02/701893/33422/1003

ECOLAB Engineering GmbH  
D-83313 Siegedorf



Anlaufgrenze 50 [l/h]\*  
lower measuring range limit

Impulszahl 70 [n/min]  
pulse rate

bei 80 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 30.08.22

Prüfer P3

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 02100

PN: **280045**



OGM PLUS 02100VCFPPKKE-G5/4I-99-0020m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 20,0 ml/pulse, 50 pulses/l

max. pump size: 210 / 250 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 02/701893/33422/1004

ECOLAB Engineering GmbH  
D-83313 Stegedorf



Anlaufgrenze 50 [l/h]\*  
lower measuring range limit

Impulszahl 71 [n/min]  
pulse rate

bei 80 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 30.08.22

Prüfer P3

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 02100

PN: **280045**

OGM PLUS 02100VCFPPKKE-G5/4I-99-0020m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 20,0 ml/pulse, 50 pulses/l

max. pump size: 210 / 250 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 02/701893/33422/1005

ECOLAB Engineering GmbH  
D-83313 Siegsdorf



Anlaufgrenze 50 [l/h]\*  
lower measuring range limit

Impulszahl 72 [n/min]  
pulse rate

bei 80 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 30.08.22

Prüfer P3

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 02100

PN: **280045**

OGM PLUS 02100VCFPPKKE-G5/4I-99-0020m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 20,0 ml/pulse, 50 pulses/l

max. pump size: 210 / 250 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 02/701893/33422/1006

ECOLAB Engineering GmbH  
D-83313 Siegedorf



Anlaufgrenze 50 [l/h]\*  
lower measuring range limit

Impulszahl 70 [n/min]  
pulse rate

bei 80 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 30.08.22

Prüfer P3

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 02100

PN: **280045**



OGM PLUS 02100VCFPPKKE-G5/4I-99-0020m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 20,0 ml/pulse, 50 pulses/l

max. pump size: 210 / 250 l/h (60/60Hz)

semiconductor junction: PNP

prod. code: 02/701893/33422/1007

ECOLAB Engineering GmbH  
D-83313 Siegsdorf



Anlaufgrenze 50 [l/h]\*  
lower measuring range limit

Impulszahl 71 [n/min]  
pulse rate

bei 80 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 30.08.22

Prüfer P3



# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 02100

PN: **280045**



OGM PLUS 02100VCFPPKKE-G5/41-99-0020m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 20,0 ml/pulse, 50 pulses/l

max. pump size: 210 / 250 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 02/701893/33422/1008

ECOLAB Engineering GmbH  
D-83313 Siegedorf



Anlaufgrenze 50 [l/h]\*  
lower measuring range limit

Impulszahl 72 [n/min]  
pulse rate

bei 80 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 30.08.22

Prüfer P3



# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 02100

PN: **200045**

OGM PLUS 02100VCFPPKKE-G5/4I-99-0020m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 20,0 ml/pulse, 50 pulses/l

max. pump size: 210 / 250 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 02/701893/33422/1009

ECOLAB Engineering GmbH  
D-83313 Siegedorf



Anlaufgrenze 50 [l/h]\*  
lower measuring range limit

Impulszahl 75 [n/min]  
pulse rate

bei 80 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 30.08.22

Prüfer P3

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 02100

PN: **280045**



OGM PLUS 02100VCFPPKKE-G5/4I-99-0020m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 20,0 ml/pulse, 50 pulses/l

max. pump size: 210 / 250 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 02/701893/33422/1010

ECOLAB Engineering GmbH  
D-83313 Siegsdorf



Anlaufgrenze 50 [l/h]\*  
lower measuring range limit

Impulszahl 70 [n/min]  
pulse rate

bei 80 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 30.08.22

Prüfer P3

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 02100

PN: **280045**



OGM PLUS 02100VCFPPKKE-G5/4l-99-0020m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 20,0 ml/pulse, 50 pulses/l

max. pump size: 210 / 250 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 02/701893/33422/1011

ECOLAB Engineering GmbH  
D-83313 Siegedorf



Anlaufgrenze 50 [l/h]\*  
lower measuring range limit

Impulszahl 71 [n/min]  
pulse rate

bei 80 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 30.08.22

Prüfer P3

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 02100

PN: **280045**



OGM PLUS 02100VCFPPKKE-G5/4I-99-0020m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 20,0 ml/pulse, 50 pulses/l

max. pump size: 210 / 250 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 02/701893/33422/1012

ECOLAB Engineering GmbH  
D-83313 Siegedorf



Anlaufgrenze 50 [l/h]\*  
lower measuring range limit

Impulszahl 73 [n/min]  
pulse rate

bei 80 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 30.08.22

Prüfer P3

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 02100

PN: **280045**

OGM PLUS 02100VCFPPKKE-G5/4I-99-0020m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 20,0 ml/pulse, 50 pulses/l

max. pump size: 210 / 250 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 02/701893/33422/1013

ECOLAB Engineering GmbH  
D-83313 Siedsdorf



Anlaufgrenze 50 [l/h]\*  
lower measuring range limit

Impulszahl 71 [n/min]  
pulse rate

bei 80 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 30.08.22

Prüfer P3

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 02100

PN: **280045**

OGM PLUS 02100VCFPPKKE-G5/4I-99-0020m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 20.0 ml/pulse, 50 pulses/l

max. pump size: 210 / 250 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 02/701899/33422/1014

ECOLAB Engineering GmbH  
D-83313 Siegedorf



Anlaufgrenze 50 [l/h]\*  
lower measuring range limit

Impulszahl 65 [n/min]  
pulse rate

bei 80 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 30.08.22

Prüfer P3

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 02100

PN: **280045**



OGM PLUS 02100VCFPPKKE-G5/4I-99-0020m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 20,0 ml/pulse, 50 pulses/l

max. pump size: 210 / 250 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 02/701893/33422/1015

ECOLAB Engineering GmbH  
D-83313 Siegsdorf



Anlaufgrenze 50 [l/h]\*  
lower measuring range limit

Impulszahl 73 [n/min]  
pulse rate

bei 80 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 30.08.22

Prüfer P3



# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 02100

PN: **280045**

OGM PLUS 02100VCFPPKKE-G5/4I-99-0020m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 20,0 ml/pulse, 50 pulses/l

max. pump size: 210 / 250 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 02/701893/33422/1016

ECOLAB Engineering GmbH  
D-83313 Siegedorf



Anlaufgrenze 50 [l/h]\*  
lower measuring range limit

Impulszahl 71 [n/min]  
pulse rate

bei 80 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 30.08.22

Prüfer P3

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 02100

PN: **280045**

OGM PLUS 02100VCFPPKKE-G5/4I-99-0020m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 20,0 ml/pulse, 50 pulses/l

max. pump size: 210 / 250 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 02/7D1893/33422/1017

ECOLAB Engineering GmbH  
D-83313 Siegedorf



Anlaufgrenze 50 [l/h]\*  
lower measuring range limit

Impulszahl 68 [n/min]  
pulse rate

bei 80 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 30.08.22

Prüfer P3

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 02100

PN: **280045**



OGM PLUS 02100VCFPPKKE-G5/4l-99-0020m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 20,0 ml/pulse, 50 pulses/l

max. pump size: 210 / 250 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 02/701893 /33422 /1018

ECOLAB Engineering GmbH  
D-83313 Siegedorf



Anlaufgrenze 50 [l/h]\*  
lower measuring range limit

Impulszahl 70 [n/min]  
pulse rate

bei 80 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 30.08.22

Prüfer P3

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 02100

PN: **280045**



OGM PLUS 02100VCFPPKKE-G5/41-99-0020m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 20,0 ml/pulse, 50 pulses/l

max. pump size: 210 / 250 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 02/701893/33422/1019

ECOLAB Engineering GmbH  
D-83313 Siegsdorf



Anlaufgrenze 50 [l/h]\*  
lower measuring range limit

Impulszahl 74 [n/min]  
pulse rate

bei 80 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 30.08.22

Prüfer P3

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 02100

PN: **280045**

OGM PLUS 02100VCFPPKKE-G5/4I-99-0020m-P

materials of construction:	PVC, FPM, PEEK, ceramic
signal output (standard):	20,0 ml/pulse, 50 pulses/l
max. pump size:	210 / 250 l/h (50/60Hz)
semiconductor junction:	PNP
prod. code:	02/701893 / 33422 / 1020

ECOLAB Engineering GmbH  
D-83313 Siegedorf



Anlaufgrenze 50 [l/h]\*  
lower measuring range limit

Impulszahl 68 [n/min]  
pulse rate

bei 80 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 29.08.22

Prüfer P3