

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 01200

PN: **280048**

OGM PLUS 01200PPEPPKKE-G1/2I-99-0005m-P

materials of construction: PP, EDPM, PEEK, ceramic

signal output (standard): 5.0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 701316/43722/1001

ECOLAB Engineering GmbH  
D-83313 Siegedorf



Anlaufgrenze 20 [l/h]\*  
lower measuring range limit

Impulszahl 118 [n/min]  
pulse rate

bei 35 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 22.09.22

Prüfer P 410

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 01200

PN: **280048**



OGM PLUS 01200PPEPPKKE-G1/2I-99-0005m-P

materials of construction: PP, EDPM, PEEK, ceramic

signal output (standard): 5.0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 701916/43722/1002

ECOLAB Engineering GmbH  
D-63313 Sieddorf



Anlaufgrenze 70 [l/h]\*  
lower measuring range limit

Impulszahl 113 [n/min]  
pulse rate



bei 35 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 22.09.22

Prüfer P 410

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 01200		
PN: <b>280048</b>		
OGM PLUS 01200PPEPPKKE-G1/2I-99-0005m-P		
materials of construction:	PP, EDPM, PEEK, ceramic	
signal output (standard):	5,0 ml/pulse, 200 pulses/l	
max. pump size:	120 / 144 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	701316/43722/1003	
ECOLAB Engineering GmbH D-83313 Siegsdorf		

Anlaufgrenze 20 [l/h]\*  
lower measuring range limit

Impulszahl 110 [n/min]  
pulse rate



bei 35 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 22.09.22

Prüfer P 410

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 01200  
PN: **290048**   
OGM PLUS 01200PPEPPKKE-G1/2I-99-0005m-P  
materials of construction: PP, EDPM, PEEK, ceramic  
signal output (standard): 5.0 ml/pulse, 200 pulses/l  
max. pump size: 120 / 144 l/h (50/60Hz)  
semiconductor junction: PNP  
prod. code: 701316 / 43722 / 1004  
ECOLAB Engineering GmbH  
D-83313 Siegedorf 

Anlaufgrenze 20 [l/h]\*  
lower measuring range limit

Impulszahl 115 [n/min]  
pulse rate

bei 35 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 22.09.22

Prüfer P 410

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 01200

PN: **280048**

OGM PLUS 01200PPEPPKKE-G1/2I-99-0005m-P

materials of construction: PP, EDPM, PEEK, ceramic

signal output (standard): 5,0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 701318/43722/1005

ECOLAB Engineering GmbH  
D-83113 Siegsdorf



Anlaufgrenze 20 [l/h]\*  
lower measuring range limit

Impulszahl 120 [n/min]  
pulse rate



bei 35 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 22.09.22

Prüfer **P 410**

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 01200		
PN: <b>280048</b>		
OGM PLUS 01200PPEPPKKE-G1/2I-99-0005m-P		
materials of construction:	PP, EDPM, PEEK, ceramic	
signal output (standard):	5,0 ml/pulse, 200 pulses/l	
max. pump size:	120 / 144 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	701316 /43722/1006	
ECOLAB Engineering GmbH D-83313 Siegsdorf		

Anlaufgrenze 20 [l/h]\*  
lower measuring range limit

Impulszahl 118 [n/min]  
pulse rate

bei 35 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 22.05.22

Prüfer P 410

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 01200

PN: **280048**

OGM PLUS 01200PPEPPKKE-G1/2I-99-0005m-P

materials of construction: PP, EDPM, PEEK, ceramic

signal output (standard): 5,0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 701318 /43722 /1.007



ECOLAB Engineering GmbH  
D-83313 Siegsdorf



Anlaufgrenze 20 [l/h]\*  
lower measuring range limit

Impulszahl 117 [n/min]  
pulse rate



bei 35 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 22.09.22

Prüfer P 410

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 01200		
PN: <b>280048</b>		
OGM PLUS 01200PPEPPKKE-G1/2I-99-0005m-P		
materials of construction:	PP, EDPM, PEEK, ceramic	
signal output (standard):	5,0 ml/pulse, 200 pulses/l	
max. pump size:	120 / 144 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	701318/43722/1008	
ECOLAB Engineering GmbH D-83313 Siegedorf		

Anlaufgrenze 20 [l/h]\*  
lower measuring range limit

Impulszahl 119 [n/min]  
pulse rate

bei 35 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 22.09.22

Prüfer P 410



# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 01200

PN: **280048**



OGM PLUS 01200PPEPPKKE-G1/2I-99-0005m-P

materials of construction: PP, EDPM, PEEK, ceramic

signal output (standard): 5,0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 701318/43722/1008

ECOLAB Engineering GmbH  
D-83313 Siegedorf



Anlaufgrenze 20 [l/h]\*  
lower measuring range limit

Impulszahl 116 [n/min]  
pulse rate



bei 35 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 22.09.22

Prüfer P 410

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 01200	
PN: <b>280048</b>	
OGM PLUS 01200PPEPPKKE-G1/2I-99-0005m-P	
materials of construction:	PP, EDPM, PEEK, ceramic
signal output (standard):	5,0 ml/pulse, 200 pulses/l
max. pump size:	120 / 144 l/h (50/60Hz)
semiconductor junction:	PNP
prod. code:	701316 /43722 /1010
ECOLAB Engineering GmbH D-83313 Slegsdorf	

Anlaufgrenze 20 [l/h]\*  
lower measuring range limit

Impulszahl 113 [n/min]  
pulse rate

bei 35 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 22.09.22

Prüfer P 410

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 01200

PN: **280048**

OGM PLUS 01200PPEPPKKE-G1/2I-99-0005m-P

materials of construction: PP, EDPM, PEEK, ceramic

signal output (standard): 5,0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 701318 / 43722 / 1011

ECOLAB Engineering GmbH  
D-83313 Siegsdorf



Anlaufgrenze 20 [l/h]\*  
lower measuring range limit

Impulszahl 118 [n/min]  
pulse rate



bei 35 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 22.09.22

Prüfer P 410

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 01200		
PN: <b>280048</b>		
OGM PLUS 01200PPEPPKKE-G1/2I-99-0005m-P		
materials of construction:	PP, EDPM, PEEK, ceramic	
signal output (standard):	5,0 ml/pulse, 200 pulses/l	
max. pump size:	120 / 144 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	701316/43722/1012	
ECOLAB Engineering GmbH D-83313 Siegedorf		

Anlaufgrenze 20 [l/h]\*  
lower measuring range limit

Impulszahl 120 [n/min]  
pulse rate



bei 35 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 22.09.22

Prüfer P 410

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 01200  
PN: **280048**   
OGM PLUS 01200PPEPPKKE-G1/2I-99-0005m-P  
materials of construction: PP, EDPM, PEEK, ceramic  
signal output (standard): 5.0 ml/pulse, 200 pulses/l  
max. pump size: 120 / 144 l/h (50/60Hz)  
semiconductor junction: PNP  
prod. code: 701316 / 43722 / 1013  
ECOLAB Engineering GmbH  
D-83313 Siegedorf 

Anlaufgrenze 20 [l/h]\*  
lower measuring range limit

Impulszahl 114 [n/min]  
pulse rate



bei 35 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 22.09.22

Prüfer P 410

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 01200		
PN: <b>280048</b>		
OGM PLUS 01200PPEPPKKE-G1/2I-99-0005m-P		
materials of construction:	PP, EDPM, PEEK, ceramic	
signal output (standard):	5.0 ml/pulse, 200 pulses/l	
max. pump size:	120 / 144 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	701316/43722/1014	
ECOLAB Engineering GmbH D-83313 Siegedorf		

Anlaufgrenze 20 [l/h]\*  
lower measuring range limit

Impulszahl 116 [n/min]  
pulse rate



bei 35 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 22.09.22

Prüfer P 410

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 01200		
PN: <b>280048</b>		
OGM PLUS 01200PPEPPKKE-G1/2I-99-0005m-P		
materials of construction:	PP, EDPM, PEEK, ceramic	
signal output (standard):	5,0 ml/pulse, 200 pulses/l	
max. pump size:	120 / 144 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	701318 / 43722 / 1015	
ECOLAB Engineering GmbH D-83313 Siegedorf		

Anlaufgrenze 20 [l/h]\*  
lower measuring range limit

Impulszahl 117 [n/min]  
pulse rate

bei 35 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 27.09.22

Prüfer P 410