

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 02100

PN: **280077**

OGM PLUS 02100VCEPPKKE-G5/4l-99-0020m-P

materials of construction: PVC, EPDM, PEEK, ceramic

signal output (standard): 20.0 ml/pulse, 50 pulses/l

max. pump size: 210 / 250 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 02/703518/43722/1001

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 50 [l/h]*
lower measuring range limit

Impulszahl 74 [n/min]
pulse rate

bei 80 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 20.09.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 02100

PN: **280077**

OGM PLUS 02100VCEPPKKE-G5/41-98-0020m-P

materials of construction: PVC, EPDM, PEEK, ceramic

signal output (standard): 20,0 ml/pulse, 50 pulses/l

max. pump size: 210 / 250 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 02/703518/43722/1002

ECOLAB Engineering GmbH
D-83313 Siegsdorf



Anlaufgrenze 50 [l/h]*
lower measuring range limit

Impulszahl 72 [n/min]
pulse rate

bei 80 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 20.09.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 02100		
PN: 280077		
OGM PLUS 02100VCEPPKKE-G5/4I-99-0020m-P		
materials of construction:	PVC, EPDM, PEEK, ceramic	
signal output (standard):	20,0 ml/pulse, 50 pulses/l	
max. pump size:	210 / 250 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	02/703516/43722/1003	
ECOLAB Engineering GmbH D-83313 Siegedorf		

Anlaufgrenze 50 [l/h]*
lower measuring range limit

Impulszahl 71 [n/min]
pulse rate

bei 80 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 20.09.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 02100		
PN: 280077		
OGM PLUS 02100VCEPPKKE-G5/4I-99-0020m-P		
materials of construction:	PVC, EPDM, PEEK, ceramic	
signal output (standard):	20,0 ml/pulse, 50 pulses/l	
max. pump size:	210 / 250 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	02/703516/43722/1004	
ECOLAB Engineering GmbH D-83313 Slegsdorf		

Anlaufgrenze 50 [l/h]*
lower measuring range limit

Impulszahl 70 [n/min]
pulse rate

bei 80 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 20.09.22

Prüfer **P 410**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 02100
PN: **280077**



OGM PLUS 02100VCEPPKKE-G5/4I-99-0020m-P
materials of construction: PVC, EPDM, PEEK, ceramic
signal output (standard): 20.0 ml/pulse, 50 pulses/l
max. pump size: 210 / 250 l/h (50/60Hz)
semiconductor junction: PNP
prod. code: 02/703516 / 43722 / 1005

ECOLAB Engineering GmbH
D-83313 Siegsdorf



Anlaufgrenze 50 [l/h]*
lower measuring range limit

Impulszahl 70 [n/min]
pulse rate

bei 80 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 20.09.22

Prüfer **P 410**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 02100		
PN: 280077		
OGM PLUS 02100VCEPPKKE-G5/41-99-0020m-P		
materials of construction:	PVC, EPDM, PEEK, ceramic	
signal output (standard):	20,0 ml/pulse, 50 pulses/l	
max. pump size:	210 / 250 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	02/703516/43722/1008	
ECOLAB Engineering GmbH D-83313 Stegedorf		

Anlaufgrenze 50 [l/h]*
lower measuring range limit

Impulszahl 76 [n/min]
pulse rate

bei 80 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 20.09.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 02100

PN: **280077**

OGM PLUS 02100VCEPPKKE-G5/4I-99-0020m-P

materials of construction: PVC, EPDM, PEEK, ceramic

signal output (standard): 20.0 ml/pulse, 50 pulses/l

max. pump size: 210 / 250 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 02/703516/43722/1007

ECOLAB Engineering GmbH
D-83313 Siegsdorf



Anlaufgrenze 50 [l/h]*
lower measuring range limit

Impulszahl 73 [n/min]
pulse rate


bei 80 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 20.09.22

Prüfer **P 410**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 02100	
PN: 280077	
OGM PLUS 02100VCEPPKKE-G5/4I-99-0020m-P	
materials of construction:	PVC, EPDM, PEEK, ceramic
signal output (standard):	20,0 ml/pulse, 50 pulses/l
max. pump size:	210 / 250 l/h (50/60Hz)
semiconductor junction:	PNP
prod. code:	02/703516/43722/1008
ECOLAB Engineering GmbH	
D-83313 Siegsdorf	

Anlaufgrenze 50 [l/h]*
lower measuring range limit

Impulszahl 69 [n/min]
pulse rate



bei 80 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 20.09.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 02100	
PN: 280077	
OGM PLUS 02100VCEPPKKE-G5/4I-99-0020m-P	
materials of construction:	PVC, EPDM, PEEK, ceramic
signal output (standard):	20.0 ml/pulse, 50 pulses/l
max. pump size:	210 / 250 l/h (50/60Hz)
semiconductor junction:	PNP
prod. code:	02/703518/43722/1009
ECOLAB Engineering GmbH	
D-63313 Siegedorf	

Anlaufgrenze 50 [l/h]*
lower measuring range limit

Impulszahl 69 [n/min]
pulse rate

bei 80 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 20.09.22

Prüfer **P 410**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 02100

PN: **280077**

OGM PLUS 02100VCEPPKKE-G5/4I-99-0020m-P

materials of construction: PVC, EPDM, PEEK, ceramic

signal output (standard): 20.0 ml/pulse, 50 pulses/l

max. pump size: 210 / 250 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 02/703516/43722/1010

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 50 [l/h]*
lower measuring range limit

Impulszahl 73 [n/min]
pulse rate

bei 80 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 20.09.22

Prüfer P 410