


Prüfprotokoll OGM^{plus}

Oval gear meter type 0.1		
PN: 280001		
OGM 001VCFPPKKEI-G1/8i-99-2600-0003		
Materials of construction:	PVC, FPM, PEEK,	
	ceramic, titan	
Signal output:	2800 pulse/l	
Max. flow rate:	3 l/h	
Prod. code:	706962 / 44922 / 1001	
ECOLAB Engineering GmbH		
D-83313 Siegsdorf		

Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 51 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 16.12.22

Prüfer Sollereder

Prüfprotokoll OGM^{plus}

Oval gear meter type 0.1

PN: **280001**

OGM 001VCFPKKEI-G1/8i-99-2600-0003

Materials of construction:

PVC, FPM, PEEK,
ceramic, titan

Signal output:

2800 pulse/l

Max. flow rate:

3 l/h

Prod. code:

706962 / 44922 / 1002

ECOLAB Engineering GmbH
D-83313 Siegsdorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 56 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 16.12.22

Prüfer Sollerseder

Prüfprotokoll OGM^{plus}

Oval gear meter type 0.1

PN: 290001

OGM 001VCFPPKKEI-G1/8i-99-2600-0003

Materials of construction:

PVC, FPM, PEEK,
ceramic, titan

Signal output:

2800 pulse/l

Max. flow rate:

3 l/h

Prod. code:

706962 / 44922 / 1003

ECOLAB Engineering GmbH
D-83313 Siegsdorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 51 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 16.12.22

Prüfer Joller edel

Prüfprotokoll OGM^{plus}

Oval gear meter type 0.1

PN: **280001**

OGM 001VCFPKKEI-G1/6i-99-2600-0003

Materials of construction:

PVC, FPM, PEEK,
ceramic, titan

Signal output:

2800 pulse/l

Max. flow rate:

3 l/h

Prod. code:

706962 / 44922 / 1004

ECOLAB Engineering GmbH

D-83313 Siegedorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 52 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 16.12.22

Prüfer Sollerseder

Prüfprotokoll OGM^{plus}

Oval gear meter type 0.1

PN: 280001

OGM 001VCFPPKKEI-G1/8i-99-2800-0003

Materials of construction:

PVC, FPM, PEEK,
ceramic, titan

Signal output:

2800 pulse/l

Max. flow rate:

3 l/h

Prod. code:

706962 / 44922 / 1005

ECOLAB Engineering GmbH
D-83313 Siegsdorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 55 [n/min]
pulse rate

bei 1,25 [l/h]*
at


* = kontinuierlicher Durchfluss / continuous flow

Datum 14.12.22

Prüfer Sollnereder

Prüfprotokoll OGM^{plus}

Oval gear meter type 0.1
PN: **280001**
OGM 001VCFPPKETI-G1/8i-99-2600-0003
Materials of construction: PVC, FPM, PEEK, ceramic, titan
Signal output: 2800 pulse/l
Max. flow rate: 3 l/h
Prod. code: 706962 / 449 22 / 1006
ECOLAB Engineering GmbH
D-63313 Siegsdorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 54 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 16.12.22

Prüfer Sollerseder

Prüfprotokoll OGM^{plus}

Oval gear meter type 0.1

PN: **280001**

OGM 001VCFPPKKEI-G1/8i-99-2800-0003

Materials of construction:

PVC, FPM, PEEK,

ceramic, titan

Signal output:

2800 pulse/l

Max. flow rate:

3 l/h

Prod. code:

706962 / 44922 / 1008

ECOLAB Engineering GmbH

D-83313 Siegsdorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 52 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 16.12.22

Prüfer Sollereder

Prüfprotokoll OGM^{plus}

Oval gear meter type 0.1

PN: 280001

OGM 001VCFPKKEKI-G1/8i-99-2600-0003

Materials of construction:

PVC, FPM, PEEK,

ceramic, titan

Signal output:

2800 pulse/l

Max. flow rate:

3 l/h

Prod. code:

706962 / 44922 / 1007

ECOLAB Engineering GmbH
D-83313 Slegsdorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 51 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 14.12.22

Prüfer Sollnereder