

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200

PN: **10240556**



OGM PLUS 01200VCEPPKKE-G3/4i-99-0005m-P

materials of construction: PVC, EPDM, PEEK, ceramic

signal output (standard): 5,0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

Year/Prod. code: 2022 / 706338 / 149 22 / 1007

ECOLAB Engineering GmbH

D-83131 Siegsdorf

Made in Germany



Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl M [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 8.12.22

Prüfer P391

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200		
PN: 10240556		
OGM PLUS 01200VCEPPKKE-G3/4i-99-0005m-P		
materials of construction:	PVC, EPDM, PEEK, ceramic	
signal output (standard):	5,0 ml/pulse, 200 pulses/l	
max. pump size:	120 / 144 l/h (50/60Hz)	
semiconductor junction:	PNP	
Year/Prod. code:	2022 / 706336 / 1 49 22 / 1006	
ECOLAB Engineering GmbH D-83133 Stegsdorf Made in Germany		

Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 112 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 8.12.22

Prüfer P391

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200

PN: **10240556**



OGM PLUS 01200VCEPPKKE-G3/4i-99-0005m-P

materials of construction: PVC, EPDM, PEEK, ceramic

signal output (standard): 5,0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

Year/Prod. code: 2022 / 706336 / 149 22 / 1005

ECOLAB Engineering GmbH

D-83313 Siegsdorf

Made in Germany



Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 112 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 8.12.22

Prüfer P391

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200

PN: **10240556**



OGM PLUS 01200VCEPPKKE-G3/4I-99-0005m-P

materials of construction: PVC, EPDM, PEEK, ceramic

signal output (standard): 5.0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

Year/Prod. code: 2022 / 706336/14922/1004

ECOLAB Engineering GmbH

D-63313 Siegsdorf

Made in Germany



Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 115 [n/min]
pulse rate



bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 8.12.22

Prüfer P391

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200		
PN: 10240556		
OGM PLUS 01200VCEPPKKE-G3/4I-99-0005m-P		
materials of construction:	PVC, EPDM, PEEK, ceramic	
signal output (standard):	5,0 ml/pulse, 200 pulses/l	
max. pump size:	120 / 144 l/h (50/60Hz)	
semiconductor junction:	PNP	
Year/Prod. code:	2022 / 708336 / 14922 / 1003	
ECOLAB Engineering GmbH D-83313 Siegsdorf Made in Germany		

Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 112 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 8.12.22

Prüfer P391

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200

PN: **10240556**

OGM PLUS 01200VCEPPKKE-G3/4I-99-0005m-P

materials of construction: PVC, EPDM, PEEK, ceramic

signal output (standard): 5,0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

Year/Prod. code: 2022 / 706336 / 14922 / 1002

ECOLAB Engineering GmbH
D-83313 Siegsdorf
Made in Germany



Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 110 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 8.12.22

Prüfer P391

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200		
PN: 10240556		
OGM PLUS 01200VCEPPKKE-G3/4I-99-0005m-P		
materials of construction:	PVC, EPDM, PEEK, ceramic	
signal output (standard):	5.0 ml/pulse, 200 pulses/l	
max. pump size:	120 / 144 l/h (50/60Hz)	
semiconductor junction:	PNP	
Year/Prod. code:	2022 / 706338 / 149 22 / 1001	
ECOLAB Engineering GmbH D-83313 Siegsdorf Made in Germany		

Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 18 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 8.12.22

Prüfer P39A