

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200

PN: **10240556**

OGM PLUS 01200VCEPPKKE-G3/4I-99-0005m-P

materials of construction: PVC, EPDM, PEEK, ceramic

signal output (standard): 5,0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

Year/Prod. code: 2023 / 719147 / 52623 / 1001

ECOLAB Engineering GmbH
D-83313 Siegsdorf
Made in Germany



Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 108 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 07 Juli 2023

Prüfer P 421

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200		
PN: 10240556		
OGM PLUS 01200VCEPPKKE-G3/4i-99-0005m-P		
materials of construction:	PVC, EPDM, PEEK, ceramic	
signal output (standard):	5,0 ml/pulse, 200 pulses/l	
max. pump size:	120 / 144 l/h (50/60Hz)	
semiconductor junction:	PNP	
Year/Prod. code:	2023 / 719147 / 52823 / 1002	
ECOLAB Engineering GmbH D-83313 Siegsdorf Made in Germany		

Anlaufgrenze _____ 20 _____ [l/h]*
lower measuring range limit

Impulszahl _____ 108 _____ [n/min]
pulse rate



bei _____ 35 _____ [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum _____ P 421 07. Juli 2023 _____

Prüfer _____

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200
PN: **10240556** 
OGM PLUS 01200VCEPPKKE-G3/4I-99-0005m-P
materials of construction: PVC, EPDM, PEEK, ceramic
signal output (standard): 5,0 ml/pulse, 200 pulses/l
max. pump size: 120 / 144 l/h (50/60Hz)
semiconductor junction: PNP
Year/Prod. code: 2023 / 719147 / 52823 / 1003
ECOLAB Engineering GmbH
D-63313 Siegedorf
Made in Germany 

Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 110 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 07. Juli 2023

Prüfer P 421

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200		
PN: 10240556		
OGM PLUS 01200VCEPPKKE-G3/4I-99-0005m-P		
materials of construction:	PVC, EPDM, PEEK, ceramic	
signal output (standard):	5,0 ml/pulse, 200 pulses/l	
max. pump size:	120 / 144 l/h (50/60Hz)	
semiconductor junction:	PNP	
Year/Prod. code:	2023 / 719147 / 52623 / 1004	
ECOLAB Engineering GmbH D-83313 Slegsdorf Made in Germany		

Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 108 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 07. Juli 2023

Prüfer P 421

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200		
PN: 10240556		
OGM PLUS 01200VCEPPKKE-G3/4I-99-0005m-P		
materials of construction:	PVC, EPDM, PEEK, ceramic	
signal output (standard):	5,0 ml/pulse, 200 pulses/l	
max. pump size:	120 / 144 l/h (50/60Hz)	
semiconductor junction:	PNP	
Year/Prod. code:	2023 / 719147 / 52623 / 1005	
ECOLAB Engineering GmbH D-83313 Slegsdorf Made in Germany		

Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 109 [n/min]
pulse rate


bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 07. Juli 2023


Prüfer P 421

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200 

PN: **10240556**
OGM PLUS 01200VCEPPKKE-G3/4I-99-0005m-P

materials of construction: PVC, EPDM, PEEK, ceramic
signal output (standard): 5,0 ml/pulse, 200 pulses/l
max. pump size: 120 / 144 l/h (50/60Hz)
semiconductor junction: PNP
Year/Prod. code: 2023 / 719147 / 52623 / 1006

ECOLAB Engineering GmbH
D-83313 Siegsdorf
Made in Germany 

Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 111 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 07. Juli 2023

Prüfer P 421

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200		
PN: 10240556		
OGM PLUS 01200VCEPPKKE-G3/4i-99-0005m-P		
materials of construction:	PVC, EPDM, PEEK, ceramic	
signal output (standard):	5,0 ml/pulse, 200 pulses/l	
max. pump size:	120 / 144 l/h (50/60Hz)	
semiconductor junction:	PNP	
Year/Prod. code:	2023 / 719147 / 52623 / 1007	
ECOLAB Engineering GmbH D-83133 Slegsdorf Made in Germany		

Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 115 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 07. Juli 2023

Prüfer P 421

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200

PN: 10240556

OGM PLUS 01200VCEPPKKE-G3/4I-99-0005m-P

materials of construction: PVC, EPDM, PEEK, ceramic

signal output (standard): 5,0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

Year/Prod. code: 2023 / 719147 / 52623 / 1008

ECOLAB Engineering GmbH
D-83313 Siegsdorf
Made in Germany



Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 113 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 07. Juli 2023

Prüfer P 421

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200		
PN: 10240556		
OGM PLUS 01200VCEPPKKE-G3/4I-99-0005m-P		
materials of construction:	PVC, EPDM, PEEK, ceramic	
signal output (standard):	5.0 ml/pulse, 200 pulses/l	
max. pump size:	120 / 144 l/h (50/60Hz)	
semiconductor junction:	PNP	
Year/Prod. code:	2023 / 719147 / 52623 / 1009	
ECOLAB Engineering GmbH D-83133 Slegsdorf Made in Germany		

Anlaufgrenze ~~10~~ 20 [l/h]*
lower measuring range limit

Impulszahl 110 [n/min]
pulse rate



bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 07. Juli 2023

Prüfer P 421

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200		
PN: 10240556		
OGM PLUS 01200VCEPPKKE-G3/4i-99-0005m-P		
materials of construction:	PVC, EPDM, PEEK, ceramic	
signal output (standard):	5,0 ml/pulse, 200 pulses/l	
max. pump size:	120 / 144 l/h (50/60Hz)	
semiconductor junction:	PNP	
Year/Prod. code:	2023 / 719147 / 52823 / 1010	
ECOLAB Engineering GmbH D-83113 Slegsdorf Made in Germany		

Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 221 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 07. Juli 2023

Prüfer P 421

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200

PN: **10240556**

OGM PLUS 01200VCEPPKKE-G3/4I-99-0005m-P

materials of construction: PVC, EPDM, PEEK, ceramic

signal output (standard): 5,0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

Year/Prod. code: 2023 / 719147 / 52623 / 1011

ECOLAB Engineering GmbH
D-83313 Siegsdorf
Made in Germany



Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 116 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

07. Juli 2023

Datum _____

Prüfer P 421

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200		
PN: 10240556		
OGM PLUS 01200VCEPPKKE-G3/4I-99-0005m-P		
materials of construction:	PVC, EPDM, PEEK, ceramic	
signal output (standard):	5.0 ml/pulse, 200 pulses/l	
max. pump size:	120 / 144 l/h (50/60Hz)	
semiconductor junction:	PNP	
Year/Prod. code:	2023 / 719147 / 52823 / 1012	
ECOLAB Engineering GmbH D-63313 Siegsdorf Made in Germany		

Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 111 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 07. Juli 2023

Prüfer P 421

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200

PN: **10240556**

OGM PLUS 01200VCEPPKKE-G3/4i-99-0005m-P

materials of construction: PVC, EPDM, PEEK, ceramic

signal output (standard): 5.0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

Year/Prod. code: 2023 / 719147 / 52823 / 1013

ECOLAB Engineering GmbH
D-83313 Slegsdorf
Made in Germany



Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 108 [n/min]
pulse rate



bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 07. Juli 2023

Prüfer P 421

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200 
PN: **10240556**
OGM PLUS 01200VCEPPKKE-G3/4I-99-0005m-P
materials of construction: PVC, EPDM, PEEK, ceramic
signal output (standard): 5.0 ml/pulse, 200 pulses/l
max. pump size: 120 / 144 l/h (50/60Hz)
semiconductor junction: PNP
Year/Prod. code: 2023 / 719147/52623/1014
ECOLAB Engineering GmbH 
D-83313 Siedsdorf
Made in Germany

Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 113 [n/min]
pulse rate

bei 25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 07. Juli 2023

Prüfer P 421

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200		
PN: 10240556		
OGM PLUS 01200VCEPPKKE-G3/4I-99-0005m-P		
materials of construction:	PVC, EPDM, PEEK, ceramic	
signal output (standard):	5.0 ml/pulse, 200 pulses/l	
max. pump size:	120 / 144 l/h (50/60Hz)	
semiconductor junction:	PNP	
Year/Prod. code:	2023 / 719147 / 52823 / 1015	
ECOLAB Engineering GmbH D-83313 Slegsdorf Made in Germany		

Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 108 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 07. Juli 2023

Prüfer P 421