

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112	
W3T262906	
OGM PLUS 00112VCFPPVKE-G1/6i-99-2600p-P	
materials of construction:	PVC, FKM HF, PVDF, ceramic
signal output (standard):	0,38 ml/pulse, 2800 pulses/l
max. pump size:	11,2 / 13,5 l/h (50/60Hz)
semiconductor junction:	PNP
prod. code:	01 / 497829 / 42722 / 1001
ECOLAB Engineering GmbH D-83313 Siegsdorf	

Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 52 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 11.07.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

W3T262906

OGM PLUS 00112VCFPPVKE-G1/8i-99-2600p-P

materials of construction: PVC, FKM HF, PVDF, ceramic

signal output (standard): 0,38 ml/pulse, 2600 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01 / 497829 / 42722 / 1002

ECOLAB Engineering GmbH
D-83313 Siegsdorf



280162



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 53 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 11.07.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112		
W3T262906		
OGM PLUS 00112VCFPPVKE-G1/8I-99-2600p-P		
materials of construction:	PVC, FKM HF, PVDF, ceramic	
signal output (standard):	0,98 ml/pulse, 2600 pulses/l	
max. pump size:	11,2 / 13,5 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	01 / 497829 / 427 22 / 1003	260162
ECOLAB Engineering GmbH		
D-83313 Siegsdorf		

Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 51 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 11.07.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112		
W3T262908		
OGM PLUS 00112VCFFPVKE-G1/8i-99-2600p-P		
materials of construction:	PVC, FKM HF, PVDF, ceramic	260162
signal output (standard):	0,39 ml/pulse, 2600 pulses/l	
max. pump size:	11,2 / 13,5 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	01 / 497829 / 427 22 / 1004	
ECOLAB Engineering GmbH D-83313 Siegsdorf		

Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 53 [n/min]
pulse rate



bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 11.07.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112		
W3T262906		
OGM PLUS 00112VCFPPVKE-G1/8i-99-2600p-P		
materials of construction:	PVC, FKM HF, PVDF, ceramic	
signal output (standard):	0,38 ml/pulse, 2600 pulses/l	
max. pump size:	11,2 / 13,5 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	01 / 497829 / 427 22 / 1005	280162
ECOLAB Engineering GmbH D-83313 Siegsdorf		

Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 50 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 11.07.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112 
W3T262906
OGM PLUS 00112VCFPPVKE-G1/8i-99-2600p-P
materials of construction: PVC, FKM HF, PVDF, ceramic
signal output (standard): 0,38 ml/pulse, 2600 pulses/l
max. pump size: 11,2 / 13,5 l/h (50/60Hz)
semiconductor junction: PNP
prod. code: 01 / 497829 / 427 22 / 1006 280162
ECOLAB Engineering GmbH 
D-83313 Siegsdorf

Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 50 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 11.07.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112		
W3T262906		
OGM PLUS 00112VCFPPVKE-G1/81-99-2600p-P		
materials of construction:	PVC, FKM HF, PVDF, ceramic	
signal output (standard):	0,36 ml/pulse, 2600 pulses/l	
max. pump size:	11,2 / 13,5 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	01 / 497829 / 42722 / 1007	260162
ECOLAB Engineering GmbH D-83313 Siegsdorf		

Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 51 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 11.07.22

Prüfer P_410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112	
W3T262906	
OGM PLUS 00112VCFPPVKE-G1/8i-99-2600p-P	
materials of construction:	PVC, FKM HF, PVDF, ceramic
signal output (standard):	0.38 ml/pulse, 2600 pulses/l
max. pump size:	11.2 / 13.5 l/h (50/60-Hz)
semiconductor junction:	PNP
prod. code:	01 / 497829 / 42722 / 1008
ECOLAB Engineering GmbH	
D-83313 Stegendorf	280162

Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 53 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 11.07.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112		
W3T262906		
OGM PLUS 00112VCFPPVKE-G1/8i-99-2600p-P		
materials of construction:	PVC, FKM HF, PVDF, ceramic	
signal output (standard):	0,38 ml/pulse, 2600 pulses/l	
max. pump size:	11,2 / 13,5 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	01 / 497829 / 42722 / 1009	280162
ECOLAB Engineering GmbH D-83313 Siegedorf		

Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 52 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 11.07.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112		
W3T262906		
OGM PLUS 00112VCFPPVKE-G1/Bi-99-2600p-P		
materials of construction:	PVC, FKM HF, PVDF, ceramic	
signal output (standard):	0,38 ml/pulse, 2600 pulses/l	
max. pump size:	11,2 / 13,5 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	01 / 497829 / 42722 / 1010	280162
ECOLAB Engineering GmbH D-63313 Sieglar		

Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 51 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 11.07.22

Prüfer P 410