

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 04500		
PN: 280046		
OGM PLUS 04500VCFPPKKE-G5/4I-99-0050m-P		
materials of construction:	PVC, FPM, PEEK, ceramic	
signal output (standard):	50.0 ml/pulse, 20 pulses/l	
max. pump size:	450 / 540 l/h (50/60-Hz)	
semiconductor junction:	PNP	
prod. code:	02/700355/43022/1001	
ECOLAB Engineering GmbH D-83313 Siegenhof		

Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 41 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 01.08.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 04500		
PN: 280046		
OGM PLUS 04500VCFPPKKE-G5/4I-99-0050m-P		
materials of construction:	PVC, FPM, PEEK, ceramic	
signal output (standard):	50.0 ml/pulse, 20 pulses/l	
max. pump size:	450 / 540 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	02/700355/43022/1002	
ECOLAB Engineering GmbH D-83313 Siegsdorf		

Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 40 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 01.08.22

Prüfer **P 410**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 04500		
PN: 280048		
OGM PLUS 04500VCFPPKKE-G5/4I-99-0050m-P		
materials of construction:	PVC, FPM, PEEK, ceramic	
signal output (standard):	50.0 ml/pulse, 20 pulses/l	
max. pump size:	450 / 540 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	02/700355/43022/1003	
ECOLAB Engineering GmbH D-83313 Siegedorf		

Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 42 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 01.08.22

Prüfer **P 410**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 04500		
PN: 280048		
OGM PLUS 04500VCFPPKKE-G5/4l-99-0050m-P		
materials of construction:	PVC, FPM, PEEK, ceramic	
signal output (standard):	50,0 ml/pulse, 20 pulses/l	
max. pump size:	450 / 540 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	02/700355 /43022 /1004	
ECOLAB Engineering GmbH D-83313 Siegsdorf		

Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 40 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 01.08.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 04500

PN: **280048**

OGM PLUS 04500VCFPPKKE-G5/4I-99-0050m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 50,0 ml/pulse, 20 pulses/l

max. pump size: 450 / 540 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 02/700955/43022/1005

ECOLAB Engineering GmbH
D-83313 Stegoderl



Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 40 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 01.08.22

Prüfer P 410