

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter type 1		
PN: 280002		
OGM 010VCFPPKKEI-G1/4i-99-0820-0030		
Materials of construction:	PVC, FPM, PEEK, ceramic, titan	
Signal output:	820 pulse/l	
Max. flow rate:	30 l/h	
Prod. code:	703631 / 139 22 / 1001	
ECOLAB Engineering GmbH		
D-83313 Siegsdorf		

Anlaufgrenze 5,5 [l/h]\*  
lower measuring range limit

Impulszahl 122 [n/min]  
pulse rate

bei 9 [l/h]\*  
at


\* = kontinuierlicher Durchfluss / continuous flow

Datum 26.09.22

Prüfer P 410

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter type 1  
PN: 280002  
OGM 010VCFPPKETI-G1/4i-99-0820-0030  
Materials of construction: PVC, FPM, PEEK, ceramic, titan  
Signal output: 820 pulse/l  
Max. flow rate: 30 l/h  
Prod. code: 703631 / 13922/1002  
ECOLAB Engineering GmbH  
D-83313 Siegsdorf



Anlaufgrenze 5,5 [l/h]\*  
lower measuring range limit

Impulszahl 123 [n/min]  
pulse rate

bei 9 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 26.09.22

Prüfer P 410