


Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 04500
PN: **280046** 
OGM PLUS 04500VCFPPKKE-G5/4I-99-0050m-P
materials of construction: PVC, FPM, PEEK, ceramic
signal output (standard): 50,0 ml/pulse, 20 pulses/l
max. pump size: 450 / 540 l/h (50/60Hz)
semiconductor junction: PNP
prod. code: 02/704944 /13822 /1001

ECOLAB Engineering GmbH
D-63313 Siegedorf 

Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 41 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 22.09.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 04500
PN: **280046** 
OGM PLUS 04500VCFPPKKE-G5/4I-99-0050m-P
materials of construction: PVC, FPM, PEEK, ceramic
signal output (standard): 50.0 ml/pulse, 20 pulses/l
max. pump size: 450 / 540 l/h (50/60Hz)
semiconductor junction: PNP
prod. code: 02/704944 / 13822 / 1002

ECOLAB Engineering GmbH 
D-83313 Siegsdorf

Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 43 [n/min]
pulse rate



bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 22.09.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 04500		
PN: 280046		
OGM PLUS 04500VCFPPKKE-G5/4i-99-0050m-P		
materials of construction:	PVC, FPM, PEEK, ceramic	
signal output (standard):	50,0 ml/pulse, 20 pulses/l	
max. pump size:	450 / 540 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	02/704944 /13822 /1003	
ECOLAB Engineering GmbH D-83313 Siegedorf		

Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 42 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 22.09.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 04500	
PN: 280048	
OGM PLUS 04500VCFPPKKE-G5/4I-99-0050m-P	
materials of construction:	PVC, FPM, PEEK, ceramic
signal output (standard):	50,0 ml/pulse, 20 pulses/l
max. pump size:	450 / 540 l/h (50/60Hz)
semiconductor junction:	PNP
prod. code:	02/704944/13622/1004
ECOLAB Engineering GmbH D-83313 Siegsdorf	

Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 40 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 22.05.22

Prüfer P_410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 04500		
PN: 280046		
OGM PLUS 04500VCFPPKKE-G5/4I-99-0050m-P		
materials of construction:	PVC, FPM, PEEK, ceramic	
signal output (standard):	50,0 ml/pulse, 20 pulses/l	
max. pump size:	450 / 540 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	02/704944/13822/1005	
ECOLAB Engineering GmbH		
D-83313 Siegedorf		

Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 41 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 22.09.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 04500

PN: **290046**

OGM PLUS 04500VCFPPKKE-G5/4I-99-0050m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 50,0 ml/pulse, 20 pulses/l

max. pump size: 450 / 540 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 02/704944 / 13822 / 1006

ECOLAB Engineering GmbH

D-83131 Siegsdorf



Anlaufgrenze 75 [l/h]*
lower measuring range limit

Impulszahl 42 [n/min]
pulse rate

bei 120 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 22.03.22

Prüfer P 410