

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**



OGM PLUS 00112VCEPVCKE-G1/8I-99-0001 m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60-Hz)

semiconductor junction: PNP

prod. code: 01/704951/24722/1065

ECOLAB Engineering GmbH
D-83313 Siegsdorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 19 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 26.11.22

Prüfer Solleroder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**



OGM PLUS 00112VCEPVCKE-G1/8I-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic
signal output (standard): 1,0 ml/pulse, 1000 pulses/l
max. pump size: 11,2 / 13,5 l/h (50/60Hz)
semiconductor junction: PNP
prod. code: 01/704951/24722/1084

ECOLAB Engineering GmbH
D-83313 Stegedorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 19 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 26.11.22

Prüfer Sollereder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**

OGM PLUS 00112VCEPVCKE-G1/8I-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1.0 ml/pulse, 1000 pulses/l

max. pump size: 11.2 / 13.5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/704951 / 24722 / 1083

ECOLAB Engineering GmbH

D-83313 Siegsdorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 19 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 26.11.22

Prüfer Sollnereder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**



OGM PLUS 00112VCEPVCKE-G1/8I-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic
signal output (standard): 1,0 ml/pulse, 1000 pulses/l
max. pump size: 11,2 / 13,5 l/h (50/60Hz)
semiconductor junction: PNP
prod. code: 01/704951 / 24722 / 1082

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 19 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 26.11.22

Prüfer Sollwender

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**

OGM PLUS 00112VCEPVCKE-G1/8I-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/704951/24722/1061

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 21 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 26.11.22

Prüfer Sollereder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**

OGM PLUS 00112VCEPVCKE-G1/8I-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/704951/24722/1080

ECOLAB Engineering GmbH
D-83313 Siegsdorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 19 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 24.11.22

Prüfer Sollgeseder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112



PN: **280151**

OGM PLUS 00112VCEPVCKE-G1/BI-99-0001m-P

materials of construction:	PVC, EPDM, PVDF, ceramic
signal output (standard):	1,0 ml/pulse, 1000 pulses/l
max. pump size:	11,2 / 13,5 l/h (50/60Hz)
semiconductor junction:	PNP
prod. code:	01/704951/24722/1059

ECOLAB Engineering GmbH
D-83313 Siegsdorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 20 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 24.11.22

Prüfer Sollaredes

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**



OGM PLUS 00112VCEPVCKE-G1/81-99-0001 m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/704951 / 24722 / 1058

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 21 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 26.11.22

Prüfer Sollereder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**

OGM PLUS 00112VCEPVCKE-G1/8I-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01 / 704951 / 24722 / 1057

ECOLAB Engineering GmbH
D-63313 Siegsdorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 21 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 26.11.22

Prüfer Sollerseder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**

OGM PLUS 00112VCEPVCKE-G1/8I-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/704951 / 24722 / 1058

ECOLAB Engineering GmbH
D-83313 Siegsdorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 20 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 26.11.22

Prüfer Sollner

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**

OGM PLUS 00112VCEPVCKE-G1/8I-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/704951 / 24722 / 1055

ECOLAB Engineering GmbH
D-83113 Siegenhof



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 19 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 26.11.22

Prüfer Sollnereder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**

OGM PLUS 00112VCEPVCKE-G1/8I-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60-Hz)

semiconductor junction: PNP

prod. code: 01/704951 / 24722 / 1054

ECOLAB Engineering GmbH
D-83313 Siegsdorf

Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 20 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 26.11.22

Prüfer Sollner

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**

OGM PLUS 00112VCEPVCKE-G1/8I-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/704951/24722/1053

ECOLAB Engineering GmbH
D-83313 Siegsdorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 19 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 26.11.22

Prüfer Sollerseder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**

OGM PLUS 00112VCEPVCKE-G1/8i-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60-Hz)

semiconductor junction: PNP

prod. code: 01/704951/24722/1052

ECOLAB Engineering GmbH
D-83133 Siegsdorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 21 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 24.11.22

Prüfer J. K. K. K.

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**

OGM PLUS 00112VCEPVCKE-G1/8I-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/704951 / 24722 / 1051

ECOLAB Engineering GmbH
D-83313 Siegsdorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 21 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 26.11.22

Prüfer Sollereder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**



OGM PLUS 00112VCEPVCKE-G1/81-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 19,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/704951/24722/1050

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 20 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 26.11.22

Prüfer Sollereder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**



OGM PLUS 00112VCEPVCKE-G1/81-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/704951/24722/1049

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 20 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 26.11.22

Prüfer Sollereder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**

OGM PLUS 00112VCEPVCKE-G1/8I-98-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/704951/24722/1048

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 19 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 24.11.22

Prüfer Sollnereder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**



OGM PLUS 00112VCEPVCKE-G1/8i-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic
signal output (standard): 1,0 ml/pulse, 1000 pulses/l
max. pump size: 11,2 / 13,5 l/h (50/60Hz)
semiconductor junction: PNP
prod. code: 01/704051 / 24722 / 1047

ECOLAB Engineering GmbH
D-83313 Siegsdorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl _____ [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum _____

Prüfer Sollner edler

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**



OGM PLUS 00112VCEPVCKE-G1/8I-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic
signal output (standard): 1,0 ml/pulse, 1000 pulses/l
max. pump size: 11,2 / 13,5 l/h (50/60Hz)
semiconductor junction: PNP
prod. code: 01/704951/24722/1048

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 19 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum _____

Prüfer Sollereder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **290151**

OGM PLUS 00112VCEPVCKE-G1/BI-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic
signal output (standard): 1,0 ml/pulse, 1000 pulses/l
max. pump size: 11,2 / 13,5 l/h (50/60Hz)
semiconductor junction: PNP
prod. code: 01/704951/24722/1045

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 21 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 26.11.22

Prüfer Sollereder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**



OGM PLUS 00112VCEPVCKE-G1/BI-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic
signal output (standard): 1,0 ml/pulse, 1000 pulses/l
max. pump size: 11,2 / 13,5 l/h (50/60Hz)
semiconductor junction: PNP
prod. code: 01/704951/24722/1044

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 19 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 24.11.22

Prüfer Sollerseder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**



OGM PLUS 00112VCEPVOKE-G1/81-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/704951/24722/1043

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 19 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 24.11.22

Prüfer Soll erdler

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**

OGM PLUS 00112VCEPVCKE-G1/8I-99-0001 m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/704951 / 24722 / 1042

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 21 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 24.11.22

Prüfer Sollereder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**

OGM PLUS 00112VCEPVCKE-G1/8I-99-0001 m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/704951/24722/1041

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 19 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 25.11.22

Prüfer Sollerseder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**

OGM PLUS 00112VCEPVCKE-G1/8I-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/704951 / 24722 / 1040

ECOLAB Engineering GmbH
D-83313 Siegsdorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 19 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 25.11.22

Prüfer Sollner

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**

OGM PLUS 00112VCEPVCKE-G1/8I-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/704951 / 24722 / 1039

ECOLAB Engineering GmbH
D-83313 Siegsdorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 20 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 25.11.22

Prüfer Sollner

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**



OGM PLUS 00112VCEPVCKE-G1/81-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/704951/24722/1038

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 08 [l/h]*
lower measuring range limit

Impulszahl 19 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 25.11.22

Prüfer Soll ercol 21

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **260151**

OGM PLUS 00112VCEPVCKE-G1/BI-99-0001m-P

materials of construction:	PVC, EPDM, PVDF, ceramic
signal output (standard):	1,0 ml/pulse, 1000 pulses/l
max. pump size:	11,2 / 13,5 l/h (50/60Hz)
semiconductor junction:	PNP
prod. code:	01 / 704951 / 24722 / 1037

ECOLAB Engineering GmbH
D-83313 Siegsdorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 21 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 25.11.22

Prüfer J. Kereses

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: 280151



OGM PLUS 00112VCEPVCKE-G1/81-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/704951 / 24722 / 1036

ECOLAB Engineering GmbH
D-83313 Siegsdorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 19 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 25.11.22

Prüfer Sollersöder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: 280151



OGM PLUS 00112VCEPVCKE-G1/81-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/704951/24722/1035

ECOLAB Engineering GmbH
D-83313 Stegendorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 20 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 25.11.22

Prüfer Sollereder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**

OGM PLUS 00112VCEPVCKE-G1/8I-99-0001 m-P

materials of construction:	PVC, EPDM, PVDF, ceramic
signal output (standard):	1,0 ml/pulse, 1000 pulses/l
max. pump size:	11,2 / 13,5 l/h (50/60Hz)
semiconductor junction:	PNP
prod. code:	01/704951 / 24722 / 1034

ECOLAB Engineering GmbH
D-83313 Siegedorf

Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 19 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 25.11.22

Prüfer Sollweide

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**

OGM PLUS 00112VCEPVCKE-G1/8i-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/704951/24722/1033

ECOLAB Engineering GmbH
D-63313 Siegedorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 20 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 25.11.22

Prüfer Sollereder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**



OGM PLUS 00112VCEPVCKE-G1/8I-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic
signal output (standard): 1,0 ml/pulse, 1000 pulses/l
max. pump size: 11,2 / 13,5 l/h (50/60Hz)
semiconductor junction: PNP
prod. code: 01/704951 / 24722 / 1032

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 20 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 25.11.22

Prüfer Pollereder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112



PN: **280151**

OGM PLUS 00112VCEPVCKE-G1/8I-99-0001m-P

materials of construction:	PVC, EPDM, PVDF, ceramic
signal output (standard):	1,0 ml/pulse, 1000 pulses/l
max. pump size:	11,2 / 13,5 l/h (50/60Hz)
semiconductor junction:	PNP
prod. code:	01/704951/24722/1031

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 21 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 25.11.22

Prüfer Sollreder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**

OGM PLUS 00112VCEPVCKE-G1/8i-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/704951 / 24722 / 1030

ECOLAB Engineering GmbH
D-63313 Siegsdorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 20 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 25.11.22

Prüfer Sollerseder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**

OGM PLUS 00112VCEPVCKE-G1/BI-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/704951/24722/1029

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 19 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 25. 11. 22

Prüfer Solgereder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**

OGM PLUS 00112VCEPVCKE-G1/8I-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic
signal output (standard): 1,0 ml/pulse, 1000 pulses/l
max. pump size: 11,2 / 13,5 l/h (50/60Hz)
semiconductor junction: PNP
prod. code: 01/704951 / 24722 / 1028

ECOLAB Engineering GmbH
D-83313 Slegsdorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 19 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 25.11.22

Prüfer Sollereder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**



OGM PLUS 00112VCEPVCKE-G1/5l-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/704951 / 24722 / 1027

ECOLAB Engineering GmbH
D-83313 Siegsdorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 19 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 25.11.22

Prüfer Soll er edler

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**

OGM PLUS 00112VCEPVCKE-G1/8I-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/704951/24722/1026

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 19 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 25.11.22

Prüfer Jallereder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**



OGM PLUS 00112VCEPVCKE-G1/8i-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/704951/24722/1025

ECOLAB Engineering GmbH
D-83313 Slegendorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 21 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 25.11.22

Prüfer Sollwreder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**

OGM PLUS 00112VCEPVCKE-G1/8I-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/704951/24722/1024

ECOLAB Engineering GmbH
D-63313 Siegedorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 20 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 25.11.22

Prüfer Sollerseder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**



OGM PLUS 00112VCEPVCKE-G1/8I-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic
signal output (standard): 1,0 ml/pulse, 1000 pulses/l
max. pump size: 11,2 / 13,5 l/h (50/60Hz)
semiconductor junction: PNP
prod. code: 01/704951/24722/1023

ECOLAB Engineering GmbH
D-83313 Siegsdorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 19 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 25.11.22

Prüfer Sollereder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **260151**



OGM PLUS 00112VCEPVCKE-G1/8I-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramio
signal output (standard): 1,0 ml/pulse, 1000 pulses/l
max. pump size: 11,2 / 13,5 l/h (50/60Hz)
semiconductor junction: PNP
prod. code: 01/704951/24722/1022

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 18 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 25.11.22

Prüfer Sollereder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**

OGM PLUS 00112VCEPVCKE-G1/81-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic
signal output (standard): 1,0 ml/pulse, 1000 pulses/l
max. pump size: 11,2 / 13,5 l/h (50/60Hz)
semiconductor junction: PNP
prod. code: 01/704951 / 24722 / 1021

ECOLAB Engineering GmbH
D-83313 Siegsdorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 20 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 25.11.22

Prüfer Sollwender

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**

OGM PLUS 00112VCEPVCKE-G1/8I-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/704951 / 24722 / 1020

ECOLAB Engineering GmbH
D-83313 Slegsdorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 21 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 25.11.22

Prüfer Sollerseder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**

OGM PLUS 00112VCEPVCKE-G1/8I-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/704951/24722/1019

ECOLAB Engineering GmbH
D-83113 Siegedorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 20 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 25.11.22

Prüfer S. Mercedes

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**

OGM PLUS 00112VCEPVCKE-G1/BI-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/704951 / 24722 / 1018

ECOLAB Engineering GmbH
D-83313 Siegedorf

Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 21 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 25.11.22

Prüfer Sollerseder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**



OGM PLUS 00112VCEPVCKE-G1/8i-99-0001 m-P

materials of construction: PVC, EPDM, PVDF, ceramic
signal output (standard): 1,0 ml/pulse, 1000 pulses/l
max. pump size: 11,2 / 13,5 l/h (50/60Hz)
semiconductor junction: PNP
prod. code: 01/704951/24722/1017

ECOLAB Engineering GmbH
D-83313 Stegendorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 21 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 25.11.22

Prüfer Sollereder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**



OGM PLUS 00112VCEPVCKE-G1/8I-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/704951/24722/1016

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 19 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 25.11.22

Prüfer Sollereder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**



OGM PLUS 00112VCEPVCKE-G1/8I-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/704951/24722/1015

ECOLAB Engineering GmbH
D-63313 Sieglar



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 19 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 25.11.22

Prüfer Sell eroder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**

OGM PLUS 00112VCEPVCKE-G1/8I-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/704951/24722/1014

ECOLAB Engineering GmbH
D-83313 Siegsdorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 19 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 25.11.22

Prüfer Sollner et al

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**



OGM PLUS 00112VCEPVCKE-G1/8I-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/704051/24722/1013

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 22 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 25.11.22

Prüfer Sollereder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**

OGM PLUS 00112VCEPVCKE-G1/BI-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/704951 / 24722/1012

ECOLAB Engineering GmbH
D-83313 Siegsdorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 21 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 25.11.22

Prüfer Sollereder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**

OGM PLUS 00112VCEPVCKE-G1/8I-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic
signal output (standard): 1,0 ml/pulse, 1000 pulses/l
max. pump size: 11,2 / 13,5 l/h (50/60Hz)
semiconductor junction: PNP
prod. code: 01/704951 / 24722 / 1011

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 20 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 25.11.22

Prüfer Joll es ed er

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**

OGM PLUS 00112VCEPVCKE-G1/8I-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/704951 / 24722 / 1010

ECOLAB Engineering GmbH
D-83313 Siegsdorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 19 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 25.11.22

Prüfer Soller edes

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**

OGM PLUS 00112VCEPVCKE-G1/8I-00-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1.0 m/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/704951 / 24722 / 1009

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 19 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 25.11.22

Prüfer Solleröder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**

OGM PLUS 00112VCEPVCKE-G1/8I-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/704951/24722/1008

ECOLAB Engineering GmbH
D-83313 Siegsdorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 19 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 25.11.22

Prüfer Soller eder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **260151**

OGM PLUS 00112VCEPVCKE-G1/8I-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/704951/24722/1007

ECOLAB Engineering GmbH
D-83313 Siegsdorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 19 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 25.11.22

Prüfer Sollerseder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**

OGM PLUS 00112VCEPVCKE-G1/8I-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/704951/24722/1006

ECOLAB Engineering GmbH
D-83313 Siegsdorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 20 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 25.11.22

Prüfer Sollweber

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**



OGM PLUS 00112VCEPVCKE-G1/81-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/704951/24722/1005

ECOLAB Engineering GmbH
D-83313 Slegsdorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 20 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 25.11.22

Prüfer Sollereder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**



OGM PLUS 00112VCEPVOKE-G1/81-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/704951/24722/1004

ECOLAB Engineering GmbH
D-83113 Siegsdorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 21 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 25.11.22

Prüfer Sollereder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**



OGM PLUS 00112VCEPVCKE-G1/8i-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic
signal output (standard): 1,0 ml/pulse, 1000 pulses/l
max. pump size: 11,2 / 13,5 l/h (50/60Hz)
semiconductor junction: PNP
prod. code: 01/704851 / 24722 / 1003

ECOLAB Engineering GmbH
D-83313 Siegsdorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 20 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 25.11.22

Prüfer Sollner edler

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**

OGM PLUS 00112VCEPVCKE-G1/BI-99-0001m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/704951 / 24722 / 1002

ECOLAB Engineering GmbH
D-83313 Stegedorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 19 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 25.11.22

Prüfer Sollereder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280151**

OGM PLUS 00112VCEPVCKE-G1/8I-99-0001 m-P

materials of construction: PVC, EPDM, PVDF, ceramic

signal output (standard): 1,0 ml/pulse, 1000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/704951/24722/1001

ECOLAB Engineering GmbH
D-83313 Siegsdorf

Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 21 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 25.11.22

Prüfer Sollersöder