


Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200 
PN: **W3T267463**
OGM PLUS 01200VCFPPKKE-G3/4I-99-0300p-P
materials of construction: PVC, FPM, PEEK, ceramic
signal output (standard): 3.33 ml/pulse, 300 pulses/l
max. pump size: 120 / 144 l/h (50/60Hz)
semiconductor junction: PNP
prod. code: 01 / 712753 / 20923 / 1001
ECOLAB Engineering GmbH 
D-63313 Slegendorf

Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 178 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28. 02. 2023

P 421

Prüfer _____

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200



PN: **W3T267463**

OGM PLUS 01200VCFPPKKE-G3/4I-99-0300P-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 3,33 ml/pulse, 300 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/712753/20923/1002

ECOLAB Engineering GmbH
D-83313 Siegsdorf



Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 175 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28.02.2023

Prüfer P 421

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200	
PN: W3T267463	
OGM PLUS 01200VCFPPKKE-G3/4I-99-0300p-P	
materials of construction:	PVC, FPM, PEEK, ceramic
signal output (standard):	3,33 ml/pulse, 300 pulses/l
max. pump size:	120 / 144 l/h (50/60Hz)
semiconductor junction:	PNP
prod. code:	01/712753 / 20923 / 1003
ECOLAB Engineering GmbH	
D-83313 Siegedorf	

Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 170 [n/min]
pulse rate



bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28.02. 2023

Prüfer P 421

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200 
PN: **W3T267463**
OGM PLUS 01200VCFPPKKE-G3/4I-99-0300p-P
materials of construction: PVC, FPM, PEEK, ceramic
signal output (standard): 3,33 ml/pulse, 300 pulses/l
max. pump size: 120 / 144 l/h (50/60Hz)
semiconductor junction: PNP
prod. code: 01/712753/20923/1004
ECOLAB Engineering GmbH
D-83313 Siegedorf 

Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 165 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28.02.2023

Prüfer P 421

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200 
PN: **W3T267463**
OGM PLUS 01200VCFPPKKE-G3/4I-99-0300p-P
materials of construction: PVC, FPM, PEEK, ceramic
signal output (standard): 3,33 ml/pulse, 300 pulses/l
max. pump size: 120 / 144 l/h (50/60Hz)
semiconductor junction: PNP
prod. code: 01/712753/20923/1005
ECOLAB Engineering GmbH
D-83313 Siegedorf 

Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 169 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28.02.2023

Prüfer P 421