



Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200		
PN: 280044		
OGM PLUS 01200VCFPPKKE-G3/4I-99-0010m-P		
materials of construction:	PVC, FPM, PEEK, ceramic	
signal output (standard):	10,0 ml/pulse, 100 pulses/l	
max. pump size:	120 / 144 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	01/497033/22422/1001	
ECLAB Engineering GmbH D-83313 Siegedorf		

Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 54 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 22.06.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200 
PN: **280044**
OGM PLUS 01200VCFPPKKE-G3/4I-00-0010m-P
materials of construction: PVC, FPM, PEEK, ceramic
signal output (standard): 10,0 ml/pulse, 100 pulses/l
max. pump size: 120 / 144 l/h (50/60Hz)
semiconductor junction: PNP
prod. code: 01/497033 / 22422 / 1002
ECOLAB Engineering GmbH
D-83313 Siegedorf 

Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 57 [n/min]
pulse rate



bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 22.06.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200		
PN: 280044		
OGM PLUS 01200VCFPPKKE-G3/4I-99-0010m-P		
materials of construction:	PVC, FPM, PEEK, ceramic	
signal output (standard):	10,0 m/pulse, 100 pulses/l	
max. pump size:	120 / 144 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	01/497033 / 22422 / 1003	
ECOLAB Engineering GmbH D-83313 Siegedorf		

Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 56 [n/min]
pulse rate

bei 35 [l/h]*
at



* = kontinuierlicher Durchfluss / continuous flow

Datum 22.06.22

Prüfer **P 410**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200
PN: **290044**
OGM PLUS 01200VCFPPKKE-G3/4I-99-0010m-P
materials of construction: PVC, FPM, PEEK, ceramic
signal output (standard): 10,0 ml/pulse, 100 pulses/l
max. pump size: 120 / 144 l/h (50/60Hz)
semiconductor junction: PNP
prod. code: 01 / 497033 / 22422 / 1004
ECOLAB Engineering GmbH
D-63313 Sieddorf



Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 58 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 22.06.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200 
PN: **260044**
OGM PLUS 01200VCFPPKKE-G3/4I-99-0010m-P
materials of construction: PVC, FPM, PEEK, ceramic
signal output (standard): 10,0 ml/pulse, 100 pulses/l
max. pump size: 120 / 144 l/h (50/60Hz)
semiconductor junction: PNP
prod. code: 01/497033 / 22422 / 1005
ECOLAB Engineering GmbH
D-83313 Siegsdorf 

Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 57 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 22.06.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200		
PN: 280044		
OGM PLUS 01200VCFPPKKE-G3/4I-99-0010m-P		
materials of construction:	PVC, FPM, PEEK, ceramic	
signal output (standard):	10.0 ml/pulse, 100 pulses/l	
max. pump size:	120 / 144 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	01/497033/22422/1008	
ECOLAB Engineering GmbH		
D-83313 Slegsdorf		

Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 60 [n/min]
pulse rate



bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 22.06.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200		
PN: 260044		
OGM PLUS 01200VCFPPKKE-G3/4I-99-0010m-P		
materials of construction:	PVC, FPM, PEEK, ceramic	
signal output (standard):	10,0 ml/pulse, 100 pulses/l	
max. pump size:	120 / 144 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	01/497033/22422/1007	
ECOLAB Engineering GmbH		
D-83313 Siedsdorf		

Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 56 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 22.06.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200		
PN: 280044		
OGM PLUS 01200VCFPPKKE-G3/4I-99-0010m-P		
materials of construction:	PVC, FPM, PEEK, ceramic	
signal output (standard):	10,0 ml/pulse, 100 pulses/l	
max. pump size:	120 / 144 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	01/497033 / 22422 / 1008	
ECOLAB Engineering GmbH		
D-83313 Siegedorf		

Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 55 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 22.06.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200

PN: **280044**



OGM PLUS 01200VCFPPKKE-G3/4I-99-0010m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 10.0 ml/pulse, 100 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/497033/22422/1009

ECOLAB Engineering GmbH

D-83313 Siegedorf



Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 57 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 22.06.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200

PN: **280044**

OGM PLUS 01200VCFPPKKE-G3/4I-99-0010m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 10.0 ml/pulse, 100 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/497033/22422/1010

ECOLAB Engineering GmbH

D-83313 Siegedorf



Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 57 [n/min]
pulse rate



bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 22.06.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200		
PN: 280044		
OGM PLUS 01200VCFPPKKE-G3/4I-99-0010m-P		
materials of construction:	PVC, FPM, PEEK, ceramic	
signal output (standard):	10,0 ml/pulse, 100 pulses/l	
max. pump size:	120 / 144 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	01/497033/22422/1011	
ECOLAB Engineering GmbH D-83313 Siegsdorf		

Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 57 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 22.06.22

Prüfer **P 410**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200
PN: **280044** 
OGM PLUS 01200VCFPPKKE-G3/4I-99-0010m-P
materials of construction: PVC, FPM, PEEK, ceramic
signal output (standard): 10.0 ml/pulse, 100 pulses/l
max. pump size: 120 / 144 l/h (50/60Hz)
semiconductor junction: PNP
prod. code: 01/497033/22422/1012
ECOLAB Engineering GmbH
D-63313 Sieglar 

Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 54 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 22.06.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200		
PN: 280044		
OGM PLUS 01200VCFPPKKE-G3/4I-99-0010m-P		
materials of construction:	PVC, FPM, PEEK, ceramic	
signal output (standard):	10,0 ml/pulse, 100 pulses/l	
max. pump size:	120 / 144 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	01 / 497033 / 22422 / 1013	
ECOLAB Engineering GmbH D-83313 Siegedorf		

Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 54 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 22.06.22

Prüfer **P 410**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200	
PN: 280044	
OGM PLUS 01200VCFPPKKE-G3/4I-99-0010m-P	
materials of construction:	PVC, FPM, PEEK, ceramic
signal output (standard):	10,0 ml/pulse, 100 pulses/l
max. pump size:	120 / 144 l/h (50/60Hz)
semiconductor junction:	PNP
prod. code:	01/497033/22422/1014
ECOLAB Engineering GmbH D-83313 Siegsdorf	

Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 56 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 22.06.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200
PN: **280044**
OGM PLUS 01200VCFPPKKE-G3/4I-99-0010m-P
materials of construction: PVC, FPM, PEEK, ceramic
signal output (standard): 10,0 ml/pulse, 100 pulses/l
max. pump size: 120 / 144 l/h (50/60Hz)
semiconductor junction: PNP
prod. code: 01/497033/22422/1015
ECOLAB Engineering GmbH
D-83313 Slegendorf

Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 57 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 22.06.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200		
PN: 280044		
OGM PLUS 01200VCFPPKKE-G3/4I-99-0010m-P		
materials of construction:	PVC, FPM, PEEK, ceramic	
signal output (standard):	10,0 ml/pulse, 100 pulses/l	
max. pump size:	120 / 144 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	01/497033/22422/1016	
ECOLAB Engineering GmbH D-83313 Siegsdorf		

Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 55 [n/min]
pulse rate



bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 22.06.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200
PN: **280044** 
OGM PLUS 01200VCFPPKKE-G3/4I-99-0010m-P
materials of construction: PVC, FPM, PEEK, ceramic
signal output (standard): 10,0 ml/pulse, 100 pulses/l
max. pump size: 120 / 144 l/h (50/60Hz)
semiconductor junction: PNP
prod. code: 01/497033 / 22422 / 1017
ECOLAB Engineering GmbH 
D-83313 Siegedorf

Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 57 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 22.06.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200

PN: **280044**

OGM PLUS 01200VCFPPKKE-G3/4I-99-0010m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 10,0 ml/pulse, 100 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/497033/22422/1018

ECOLAB Engineering GmbH

D-83313 Siegsdorf



Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 54 [n/min]
pulse rate



bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 22.06.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200	
PN: 280044	
OGM PLUS 01200VCFPPKKE-G3/4I-99-0010m-P	
materials of construction:	PVC, FPM, PEEK, ceramic
signal output (standard):	10,0 ml/pulse, 100 pulses/l
max. pump size:	120 / 144 l/h (50/60Hz)
semiconductor junction:	PNP
prod. code:	01/497033 / 22422 / 1019
ECOLAB Engineering GmbH	
D-83313 Siegsdorf	

Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 54 [n/min]
pulse rate


bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 22.06.22

Prüfer **P 410**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200		
PN: 280044		
OGM PLUS 01200VCFPPKKE-G3/4I-99-0010m-P		
materials of construction:	PVC, FPM, PEEK, ceramic	
signal output (standard):	10,0 ml/pulse, 100 pulses/l	
max. pump size:	120 / 144 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	01 / 497033 / 22422 / 1020	
ECOLAB Engineering GmbH		
D-83313 Siegsdorf		

Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 59 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 22.06.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200
PN: **260044**
OGM PLUS 01200VCFPPKKE-G3/4I-99-0010m-P
materials of construction: PVC, FPM, PEEK, ceramic
signal output (standard): 10,0 ml/pulse, 100 pulses/l
max. pump size: 120 / 144 l/h (50/60Hz)
semiconductor junction: PNP
prod. code: 01/497033/22422/1021
ECOLAB Engineering GmbH
D-83313 Siegedorf

Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 54 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 22.06.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200		
PN: 280044		
OGM PLUS 01200VCFPPKKE-G3/4I-00-0010m-P		
materials of construction:	PVC, FPM, PEEK, ceramic	
signal output (standard):	10,0 ml/pulse, 100 pulses/l	
max. pump size:	120 / 144 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	01/497033/22422/1022	
ECOLAB Engineering GmbH D-83313 Siegedorf		

Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 56 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 22.06.22

Prüfer **P 410**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200		
PN: 280044		
OGM PLUS 01200VCFPPKKE-G3/4I-99-0010m-P		
materials of construction:	PVC, FPM, PEEK, ceramic	
signal output (standard):	10,0 ml/pulse, 100 pulses/l	
max. pump size:	120 / 144 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	01/497033/22422/1023	
ECOLAB Engineering GmbH		
D-83313 Siegsdorf		

Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 56 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 22.06.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200		
PN: 280044		
OGM PLUS 01200VCFPPKKE-G3/4I-99-0010m-P		
materials of construction:	PVC, FPM, PEEK, ceramic	
signal output (standard):	10,0 ml/pulse, 100 pulses/l	
max. pump size:	120 / 144 l/h (60/60Hz)	
semiconductor junction:	PNP	
prod. code:	01/497033/22422/1024	
ECOLAB Engineering GmbH D-83313 Siegedorf		

Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 57 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 22.06.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200		
PN: 280044		
OGM PLUS 01200VCFPPKKE-G3/4I-99-0010m-P		
materials of construction:	PVC, FPM, PEEK, ceramic	
signal output (standard):	10,0 ml/pulse, 100 pulses/l	
max. pump size:	120 / 144 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	01/497033 / 22422 / 1025	
ECOLAB Engineering GmbH		
D-83313 Blegsdorf		

Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 55 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 22.06.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200		
PN: 280044		
OGM PLUS 01200VCFPPKKE-G3/4I-99-0010m-P		
materials of construction:	PVC, FPM, PEEK, ceramic	
signal output (standard):	10,0 ml/pulse, 100 pulses/l	
max. pump size:	120 / 144 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	01/497033/22422/1026	
ECOLAB Engineering GmbH D-83313 Siegsdorf		

Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 57 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 22.06.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200		
PN: 280044		
OGM PLUS 01200VCFPPKKE-G3/4I-99-0010m-P		
materials of construction:	PVC, FPM, PEEK, ceramic	
signal output (standard):	10,0 ml/pulse, 100 pulses/l	
max. pump size:	120 / 144 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	01 / 497033 / 22422 / 1027	
ECOLAB Engineering GmbH D-83313 Siegsdorf		

Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 58 [n/min]
pulse rate


bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 22.06.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200		
PN: 280044		
OGM PLUS 01200VCFPPKKE-G3/4I-99-0010m-P		
materials of construction:	PVC, FPM, PEEK, ceramic	
signal output (standard):	10,0 ml/pulse, 100 pulses/l	
max. pump size:	120 / 144 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	01/497033/22422/1028	
ECOLAB Engineering GmbH D-83313 Slegsdorf		

Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 59 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 22.06.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200	
PN: 280044	
OGM PLUS 01200VCFPPKKE-G3/4I-99-0010m-P	
materials of construction:	PVC, FPM, PEEK, ceramic
signal output (standard):	10,0 ml/pulse, 100 pulses/l
max. pump size:	120 / 144 l/h (50/60Hz)
semiconductor junction:	PNP
prod. code:	01/497033/22422/1029
ECOLAB Engineering GmbH	
D-83313 Siegedorf	

Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 56 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 22.06.22

Prüfer **P 410**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200
PN: **280044** 
OGM PLUS 01200VCFPPKKE-G3/4I-99-0010m-P
materials of construction: PVC, FPM, PEEK, ceramic
signal output (standard): 10,0 ml/pulse, 100 pulses/l
max. pump size: 120 / 144 l/h (50/60Hz)
semiconductor junction: PNP
prod. code: 01/497033/22422/1030
ECOLAB Engineering GmbH
D-83313 Siegedorf 

Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 54 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 22.06.22

Prüfer **P 410**

Prüfprotokoll OGM^{plus}



Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 57 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 22.06.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200		
PN: 280044		
OGM PLUS 01200VCFPPKKE-G3/4I-99-0010m-P		
materials of construction:	PVC, FPM, PEEK, ceramic	
signal output (standard):	10,0 ml/pulse, 100 pulses/l	
max. pump size:	120 / 144 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	01/497033/22422/1032	
ECCLAB Engineering GmbH D-83313 Stegedorf		

Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 59 [n/min]
pulse rate



bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 22.06.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200		
PN: 280044		
OGM PLUS 01200VCFPPKKE-G3/4I-99-0010m-P		
materials of construction:	PVC, FPM, PEEK, ceramic	
signal output (standard):	10,0 ml/pulse, 100 pulses/l	
max. pump size:	120 / 144 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	01/497033/22422/1033	
ECOLAB Engineering GmbH D-83133 Siegedorf		

Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 56 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 22.06.22

Prüfer **P 410**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200

PN: **280044**

OGM PLUS 01200VCFPPKKE-G3/4I-99-0010m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 10,0 ml/pulse, 100 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/497033/22422/1034

ECOLAB Engineering GmbH

D-83313 Siegedorf



Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 56 [n/min]
pulse rate



bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 22.06.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200		
PN: 280044		
OGM PLUS 01200VCFPPKKE-G3/4I-99-0010m-P		
materials of construction:	PVC, FPM, PEEK, ceramic	
signal output (standard):	10,0 ml/pulse, 100 pulses/l	
max. pump size:	120 / 144 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	01/497033/22422/1035	
ECOLAB Engineering GmbH D-83313 Siegedorf		

Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 56 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 22.06.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200

PN: **280044**



OGM PLUS 01200VCFPPKKE-G3/4I-99-0010m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 10.0 ml/pulse, 100 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/497033/22422/1038

ECOLAB Engineering GmbH

D-83313 Siegedorf



Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 55 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 22.06.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200

PN: **260044**



OGM PLUS 01200VCFPPKKE-G3/4I-99-0010m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 10,0 ml/pulse, 100 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/497033 / 22422 / 1037

ECOLAB Engineering GmbH

D-83313 Siegedorf



Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 57 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 22.06.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200		
PN: 280044		
OGM PLUS 01200VCFPPKKE-G3/4I-99-0010m-P		
materials of construction:	PVC, FPM, PEEK, ceramic	
signal output (standard):	10,0 ml/pulse, 100 pulses/l	
max. pump size:	120 / 144 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	01/497033 / 22422 / 1038	
ECOLAB Engineering GmbH D-83313 Siegsdorf		

Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 55 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 22.06.22

Prüfer **P 410**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200	
PN: 280044	
OGM PLUS 01200VCFPPKKE-G3/4I-99-0010m-P	
materials of construction:	PVC, FPM, PEEK, ceramic
signal output (standard):	10,0 ml/pulse, 100 pulses/l
max. pump size:	120 / 144 l/h (50/60Hz)
semiconductor junction:	PNP
prod. code:	01/497033/22422/1039
ECOLAB Engineering GmbH D-83313 Siegsdorf	

Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 54 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 22.06.22

Prüfer **P 410**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200

PN: **280044**

OGM PLUS 01200VCFPPKKE-G3/4I-99-0010m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 10,0 ml/pulse, 100 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/497033/22422/1040

ECOLAB Engineering GmbH

D-63313 Siegedorf



Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 56 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 22.06.22

Prüfer **P 410**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200		
PN: 280044		
OGM PLUS 01200VCFPPKKE-G3/4I-99-0010m-P		
materials of construction:	PVC, FPM, PEEK, ceramic	
signal output (standard):	10,0 ml/pulse, 100 pulses/l	
max. pump size:	120 / 144 l/h (50/60-Hz)	
semiconductor junction:	PNP	
prod. code:	01/497033/22422/1041	
ECOLAB Engineering GmbH D-83313 Siegedorf		

Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 57 [n/min]
pulse rate



bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 22.06.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200		
PN: 280044		
OGM PLUS 01200VCFPPKKE-G3/4I-99-0010m-P		
materials of construction:	PVC, FPM, PEEK, ceramic	
signal output (standard):	10,0 ml/pulse, 100 pulses/l	
max. pump size:	120 / 144 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	01 / 497033 / 22422 / 1042	
ECOLAB Engineering GmbH D-83313 Siegedorf		

Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 56 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 22.06.22

Prüfer **P 410**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200		
PN: 280044		
OGM PLUS 01200VCFPPKKE-G3/4I-99-0010m-P		
materials of construction:	PVC, FPM, PEEK, ceramic	
signal output (standard):	10,0 ml/pulse, 100 pulses/l	
max. pump size:	120 / 144 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	01/497033/22422/1043	
ECOLAB Engineering GmbH		
D-83313 Siegsdorf		

Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 55 [n/min]
pulse rate

bei 35 [l/h]*
at



* = kontinuierlicher Durchfluss / continuous flow

Datum 22.06.22

Prüfer **P 410**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200
PN: **280044**
OGM PLUS 01200VCFPPKKE-G3/4I-99-0010m-P
materials of construction: PVC, FPM, PEEK, ceramic
signal output (standard): 10,0 ml/pulse, 100 pulses/l
max. pump size: 120 / 144 l/h (50/60Hz)
semiconductor junction: PNP
prod. code: 01/497033/22422/1044
ECOLAB Engineering GmbH
D-83313 Siegsdorf



Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 57 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 21.06.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200		
PN: 280044		
OGM PLUS 01200VCFPPKKE-G3/4i-99-0010m-P		
materials of construction:	PVC, FPM, PEEK, ceramic	
signal output (standard):	10,0 ml/pulse, 100 pulses/l	
max. pump size:	120 / 144 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	01/497033/22422/1045	
ECOLAB Engineering GmbH D-83313 Slegsdorf		

Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 55 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 21.06.22

Prüfer **P 410**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200	
PN: 280044	
OGM PLUS 01200VCFPPKKE-G3/4I-99-0010m-P	
materials of construction:	PVC, FPM, PEEK, ceramic
signal output (standard):	10,0 ml/pulse, 100 pulses/l
max. pump size:	120 / 144 l/h (50/60Hz)
semiconductor junction:	PNP
prod. code:	01/497033/22422/1046
ECOLAB Engineering GmbH D-83313 Siegedorf	

Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 56 [n/min]
pulse rate



bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 21.06.22

Prüfer **P 410**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200		
PN: 280044		
OGM PLUS 01200VCFPPKKE-G3/4I-99-0010m-P		
materials of construction:	PVC, FPM, PEEK, ceramic	
signal output (standard):	10,0 ml/pulse, 100 pulses/l	
max. pump size:	120 / 144 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	01/497033 / 22422 / 1047	
ECOLAB Engineering GmbH D-83313 Siegedorf		

Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 59 [n/min]
pulse rate



bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 21.06.22

Prüfer **P 410**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200		
PN: 280044		
OGM PLUS 01200VCFPPKE-G3/4I-99-0010m-P		
materials of construction:	PVC, FPM, PEEK, ceramic	
signal output (standard):	10,0 ml/pulse, 100 pulses/l	
max. pump size:	120 / 144 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	01/497033/22422/1048	
ECOLAB Engineering GmbH		
D-83313 Siegedorf		

Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 57 [n/min]
pulse rate


bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 21.06.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200		
PN: 280044		
OGM PLUS 01200VCFPPKKE-G3/4I-99-0010m-P		
materials of construction:	PVC, FPM, PEEK, ceramic	
signal output (standard):	10,0 ml/pulse, 100 pulses/l	
max. pump size:	120 / 144 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	01/497033 / 22422 / 1049	
ECOLAB Engineering GmbH D-83313 Siegedorf		

Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 60 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 21.06.22

Prüfer **P 410**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 01200		
PN: 280044		
OGM PLUS 01200VCFPPKKE-G3/4I-99-0010m-P		
materials of construction:	PVC, FPM, PEEK, ceramic	
signal output (standard):	10,0 ml/pulse, 100 pulses/l	
max. pump size:	120 / 144 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	01/497033/22422/1050	
ECOLAB Engineering GmbH D-83313 Siegedorf		

Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 54 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 21.06.22

Prüfer **P 410**