


Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112 
PN: **280140**
OGM PLUS 00112PVFPPVKE-G1/81-99-00005m-P
materials of construction: PVDF, FPM, PVDF, ceramic
signal output (standard): 0,5 ml/pulse, 2000 pulses/l
max. pump size: 11,2 / 13,5 l/h (50/60Hz)
semiconductor junction: PNP
prod. code: 01/701854/43822/1001
ECOLAB Engineering GmbH 
D-83313 Siegsdorf

Anlaufgrenze 0,80 [l/h]*
lower measuring range limit

Impulszahl 38 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 23.12.22

Prüfer P391

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280140**

OGM PLUS 00112PVFPPVKE-G1/8I-99-00005m-P

materials of construction: PVDF, FPM, PVDF, ceramic

signal output (standard): 0,5 ml/pulse, 2000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/701854/43822/1002

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 0,80 [l/h]*
lower measuring range limit

Impulszahl 39 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 23.12.22

Prüfer P391

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280140**

OGM PLUS 00112PVFPPVKE-G1/8I-99-00005m-P

materials of construction: PVDF, FPM, PVDF, ceramic

signal output (standard): 0,5 ml/pulse, 2000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/701854/43822/1003

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 0,80 [l/h]*
lower measuring range limit

Impulszahl 39 [n/min]
pulse rate


bei 1,25 [l/h]*
at


* = kontinuierlicher Durchfluss / continuous flow

Datum 23.12.22

Prüfer P391

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112 
PN: **280140**
OGM PLUS 00112PVFPPVKE-G1/6I-99-00005m-P
materials of construction: PVDF, FPM, PVDF, ceramic
signal output (standard): 0,5 ml/pulse, 2000 pulses/l
max. pump size: 11,2 / 13,5 l/h (50/60-Hz)
semiconductor junction: PNP
prod. code: 01/701854/43622/1004

ECOLAB Engineering GmbH 
D-83313 Siegedorf

Anlaufgrenze 0,80 [l/h]*
lower measuring range limit

Impulszahl 39 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 23.12.22

Prüfer P391

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280140**



OGM PLUS 00112PVFPPVKE-G1/8I-99-00005m-P

materials of construction: PVDF, FPM, PVDF, ceramic

signal output (standard): 0,5 ml/pulse, 2000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01 / 701854 / 43822 / 1005

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 0,80 [l/h]*
lower measuring range limit

Impulszahl 41 [n/min]
pulse rate



bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 23.12.22

Prüfer P391

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112	
PN: 280140	
OGM PLUS 00112PVFPPVKE-G1/81-88-00005m-P	
materials of construction:	PVDF, FPM, PVDF, ceramic
signal output (standard):	0.5 ml/pulse, 2000 pulses/l
max. pump size:	11.2 / 13.5 l/h (50/60Hz)
semiconductor junction:	PNP
prod. code:	01/701854/43622/1006
ECOLAB Engineering GmbH D-83313 Siegedorf	

Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 38 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 23.12.22

Prüfer P39A